

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE



Customer Name & Address:	M/S. Anusakthi Micro Lab, 5/118 N, KMS Complex, Salem Service Road, Bus Stop, Lakshmi Nagar, Erode - 638 316.	Certificate No:	AMS/LAB/2889697
		Issuing Date:	14/08/2021
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Centrifuge	Id. No:	ASM/LAB/CEN/001
Make:	Electra	Department:	Laboratory
Model:	Junior Lab	Calibrated On:	13/08/2021
Sl. No:	125045	Calibration Due Suggested:	12/08/2022
Type:	Digital	DUC Condition on Receipt:	OK
Range & Least Count:	0 - 4000RPM / LC - 1		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Digital Tachometer	Mextech / DT-2234C	PC/01/21/ATL/0339-02	22/01/2022
2	Thermo Hygrometer	Beetech/088	PC/01/21/ATL/0194-01	22/01/2022

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL.No.	DUC Reading (RPM)	Standard Reading (RPM)	Correction (RPM)
1	1000	1001	+001
2	1500	1509	+009
3	2000	2014	+014
4	3000	3245	+245
5	4000	4350	+350

Allowable Tolerance: User Defined



Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)

B. S. S.



Approved by:
(Technical Manager)

B. S. S.