

## LAB MONTHLY SUMMARY



Lab Name **CELL TECH CLINICAL LABORATORY** Lab No **6167**  
 Month **November** Year **2021**  
 Constituent Group **Chemistry II**



Date of Result Entered : 23/11/2021

Date of Report Published : 03/12/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	147	146.13	8.85	12.93	138.1 mg/dl	-0.62	2.13
2	UREA II	Urease UV / GLDH II	Mindray	121	43.41	11.21	4.87	41 mg/dl	-0.50	0.88
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	92	9.37	12.27	1.15	8.5 mg/dl	-0.76	0.24
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Mindray	131	2.30	19.68	0.45	2.15 mg/dl	-0.33	0.08
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	137	5.30	8.61	0.46	5.6 g/dl	0.66	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	137	3.24	9.35	0.30	3.6 g/dl	1.19	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	132	8.29	10.20	0.84	8.8 mg/dl	0.60	0.15
8	CHOLESTROL II	CHOD-PAP II	Mindray	142	109.67	8.11	8.89	106.1 mg/dl	-0.40	1.49
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	142	159.26	10.15	16.17	148.2 mg/dl	-0.68	2.71

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinq@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*