

## LAB MONTHLY SUMMARY



Lab Name **CELL TECH CLINICAL LABORATORY** Lab No **6167**  
 Month **September** Year **2021**  
 Constituent Group **Chemistry II**



Date of Result Entered : 22/09/2021

Date of Report Published : 05/10/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	143	403.99	11.17	45.12	348.4 mg/dl	-1.23	7.55
2	UREA II	Urease UV / GLDH II	Mindray	112	91.16	8.16	7.44	90.2 mg/dl	-0.13	1.41
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	87	3.35	10.93	0.37	3.4 mg/dl	0.14	0.08
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Mindray	125	2.12	13.90	0.30	2.2 mg/dl	0.27	0.05
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	121	5.13	7.50	0.38	5.6 g/dl	1.22	0.07
6	ALBUMIN II	BCG - Colorimetric II	Mindray	124	3.21	8.83	0.28	3.3 g/dl	0.32	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	122	8.39	12.38	1.04	9.38 mg/dl	0.95	0.19
8	CHOLESTROL II	CHOD-PAP II	Mindray	138	113.82	10.26	11.68	133 mg/dl	1.64	1.99
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	126	230.68	9.14	21.09	213 mg/dl	-0.84	3.76

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinq@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*