

LAB MONTHLY SUMMARY



Lab Name **CELL TECH CLINICAL LABORATORY** Lab No **6167**
 Month **July** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 23/07/2021

Date of Report Published : 02/08/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	142	151.78	7.48	11.36	141 mg/dl	-0.95	1.91
2	UREA II	Urease UV / GLDH II	Mindray	119	31.60	13.28	4.20	36 mg/dl	1.05	0.77
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	94	2.80	10.45	0.29	2.5 mg/dl	-1.03	0.06
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	125	2.92	12.35	0.36	2.6 mg/dl	-0.89	0.06
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	125	5.39	8.15	0.44	4.8 g/dl	-1.34	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	135	3.35	9.01	0.30	3.5 g/dl	0.50	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	130	4.85	19.12	0.93	4.8 mg/dl	-0.05	0.16
8	CHOLESTROL II	CHOD-PAP II	Mindray	144	113.36	10.13	11.48	116 mg/dl	0.23	1.91
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	136	175.27	10.64	18.64	147 mg/dl	-1.52	3.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinq@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****