

RML – Quality Assurance Program (RML – QAP)



HEMATOLOGY

ALL METHOD REPORT

Lab Code: 1097

Cycle-10/2021

Round -5

Date: 02/10/2021

Complete Blood Count (CBC)

Parameters	No. of Participants	Group Mean	Standard deviation (SD)	Uncertainty of Assign Values	Range (± 2 SD)	Your Value	Standard Deviation Index(SDI)
Hb gm/dl	153	11.6	0.4	0.04	10.7-12.4	11.2	-0.9
WBC $\times 10^3/\mu\text{l}$	151	10.2	1.8	0.19	6.5-13.9	10.3	0.1
RBC $\times 10^6/\mu\text{l}$	154	3.9	0.1	0.01	3.66-4.22	3.72	-1.6
Hct%	154	34.1	2.4	0.25	29.2-39.0	32.4	-0.7
MCV fl.	154	86.4	4.6	0.47	77.2-95.6	87.2	0.2
MCH pg.	154	29.1	1.0	0.10	27.4-31.4	30.1	0.7
MCHC gm/dl	154	34.1	2.4	0.24	29.3-38.9	34.5	0.2
Platelet $\times 10^3/\mu\text{l}$	154	266.8	23.9	2.40	213.1-308.5	242	-0.8

Interpretation of SDI:

SDI Value(+/-)	0 - 0.5	0.6 - 0.9	1.0 - 2.0	2.1 - 2.9	≥ 3
Interpretation	Excellent Performance	Good Performance	Acceptable Performance	Marginal Performance Need Improvement	Unacceptable Performance Needs Urgent action

Peripheral Blood Smear(PBS):

	Your Result	Consensus Result
DLC	P-2, L-97, M-1	L84.0-98.0, Prol2.0-20.6, Pl1.5-6.8, M1.0-2.3, E1.0-1.0
Morphology	Many smudge cells and few prolymphocytes seen. RBC: Normocytic Normochromic RBC, Severe lymphocytosis with presence of few prolymphocytes. Moderate thrombocytopenia	Δ Smudge cells (97/126) Δ Normocytic/Normochromic (89/126) Δ Thrombocytopenia (72/126) Δ Lymphocytosis (62/126) Δ Leucocytosis (36/126)
Diagnosis	Chronic Lymphoproliferative Disorder (CLL)	Chronic Lymphocytic Leukemia (CLL)/ Chronic Lymphoproliferative Disorder

Legends	(*) Excluded From Group Mean	(#) Not Reported	(#) Late Result Submission	(\$) Reported in other Unit
---------	------------------------------	------------------	----------------------------	-----------------------------

Chief Coordinator

Dr. Sanjay Mehrotra

Checked By: *[Signature]*

Programme Director

Dr. Bandana Mehrotra

End of Report


Page 1 of 5



Address: B-171, Nirala Nagar, Lucknow - 226 020. Ph : 4034100-130 (30 Lines), 4077180, 2788444 Fax : (0522)2786555

Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence



RML – Quality Assurance Program (RML – QAP)



HEMATOLOGY METHOD WISE REPORT

Cycle-10/2021

Round -5

Lab Code: 1097

Date: 02/10/2021

Note: Your lab is not the part of Method Group

Complete Blood Count (CBC)

Parameters	Method Group	No.of Participants	Group Mean	Standard deviation (SD)	Uncertainty of Assign Values	Range (± 2 SD)	Your Value	Standard Deviation Index(SDI)
Hb gm/dl	Photometric	51	11.6	0.4	0.07	10.8-12.4	-	-
WBC $\times 10^3/\mu\text{l}$	Electrical impedance	50	10.7	0.7	0.12	9.3-12.2	-	-
RBC $\times 10^6/\mu\text{l}$	Electrical impedance	57	3.9	0.1	0.02	3.7-4.2	-	-
Hct%	Calculated	49	34.5	2.6	0.47	29.2-39.8	-	-
MCV fl.	Electrical impedance	34	86.8	4.4	0.93	78.1-95.5	-	-
MCH pg.	Calculated	62	29.6	0.7	0.11	28.1-31.0	-	-
MCHC gm/dl	Calculated	63	34.2	2.5	0.39	29.1-39.1	-	-
Platelet $\times 10^3/\mu\text{l}$	Electrical impedance	53	258.3	23.9	4.10	210.5-306.1	-	-

Interpretation of SDI:

SDI Value(+/-)	0 - 0.5	0.6 - 0.9	1.0 - 2.0	2.1 - 2.9	≥ 3
Interpretation	Excellent Performance	Good Performance	Acceptable Performance	Marginal Performance Need Improvement	Unacceptable Performance Needs Urgent action

Legends	(*) Excluded From Group Mean	(#) Not Reported	(%) Late Result Submission	(\$) Reported in other Unit
---------	------------------------------	------------------	----------------------------	-----------------------------

Chief Coordinator

Dr.Sanjay Mehrotra

Checked By

Programme Director

Dr.Bandana Mehrotra

“End of Report”

Page 2 of 3



Address: B-171, Nirala Nagar, Lucknow - 226 020. Ph. : 4034100-130 (30 Lines), 4077180, 2788444 Fax: (0522)2788555

Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence