



RML – Quality Assurance Program (RML – QAP)



Urine Routine Analysis

ALL METHOD REPORT

Lab Code: 1097

Cycle - 10/2021
Round No - 9

Date: 02/10/2021

Parameters	Units	No. of Participants	Consensus		Consensus Result	Your Value	Remark
			Number	%			
Specific Gravity	-	164	141	86%	1.010 - 1.020	1.020	Within consensus
pH	-	163	134	82%	6.0 - 7.0	6.0	Within consensus
Leukocytes	/ μ l	159	154	97%	Negative	Negative	Within consensus
Nitrite	-	155	151	97%	Negative	Negative	Within consensus
Protein	mg/dl	165	159	96%	Negative	Negative	Within consensus
Glucose	mg/dl	165	161	98%	Negative	Negative	Within consensus
Ketones	-	165	161	98%	Negative	Negative	Within consensus
Urobilinogen	mg/dl	163	160	98%	Negative	Negative	Within consensus
Bilirubin	mg/dl	164	161	98%	Negative	Negative	Within consensus
Blood (Hemoglobin)	/ μ l	163	157	96%	Negative	Negative	Within consensus

Legends	\geq 60% Consensus	(*) Outside Consensus	(@) No Consensus	(.) Not Reported	(#) Late Result Submission
---------	----------------------	-----------------------	------------------	------------------	----------------------------

Chief Coordinator

Dr. Sanjay Mehrotra

Checked By

Programme Director

Dr. Bandana Mehrotra

Page 1 of 4



Address: B-171, Nirala Nagar, Lucknow - 226 020. Ph. : 4034100-130 (30 Lines), 4077180, 2788444 Fax : (0522)2788555

Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence