



Beckman Coulter India Pvt. Ltd.  
3<sup>rd</sup> floor, B-wing,  
Art Guild House  
Phoenix Market City  
LBS Road, Kurla (W)  
Mumbai- 400 070  
Ph: 022 30805000

Date: 16<sup>th</sup> July 2021

### Calibration Certificate

Analyzer Model: Beckman coulter AU480  
Instrument Serial no.: 2016014758  
Installed at: MULJIBHAI PATEL UROLOGICAL HOSPITAL  
Calibrated on: 16<sup>th</sup> July 2021  
Next calibration due: 15<sup>th</sup> July 2022

#### Photometric Check:

Requirement: All O.D. should be less than 1.7 units.

Photometric check is OK.

#### Incubation temperature:

Range: 37 +/- 0.5

On display the temperature is 37.0 and is stable.

We hereby confirm that Beckman Coulter AU480 has been successfully calibrated and passed All quality check and is working in good condition.

For, Beckman Coulter India Pvt. Ltd.

A black and white photograph of a handwritten signature, appearing to read "Anil Munjani", written over a dark, textured background.

Anil Munjani  
Field Service Engineer



Beckman Coulter India Pvt. Ltd.  
3<sup>rd</sup> floor, B-wing,  
Art Guild House  
Phoenix Market City  
LBS Road, Kurla (W)  
Mumbai- 400 070  
Ph: 022 30805000

Date: 16<sup>th</sup> July 2021

### Calibration Certificate

This is to certify that AU480 (SR. NO. 2016014758) installed at MULJIBHAI PATEL UROLOGICAL HOSPITAL, is calibrated as per the recommended procedures. The system parameters as mentioned below were verified and necessary calibrations done as applicable.

#### System temperatures

Temperature Sensor	Temperature (°C)	Operating range (°C)
Reagent Storage	8.0	4.00 to 12.00
Incubator Wheel	37.00	36.50 to 37.50

#### Power Supply Units

Power Supply	Measured (V)	Tolerances (V)
+24V	24.05	1.2V
+12V	12.01	0.4V
+5V	5.00	0.2V
-5.2V	-5.21	N.A.
-12V	-12.01	0.4V

All the parameters are within the specified Limits.

For, Beckman Coulter India Pvt. Ltd.

Anil Munjani  
Field Service Engineer