

LAB MONTHLY SUMMARY



Lab Name **PREM CLINICAL LAB** Lab No **13171**
 Month **November** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 22/11/2021

Date of Report Published : 03/12/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Biosystems	275	145.00	8.96	12.99	138 mg/dl	-0.54	1.57
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	844	43.73	12.85	5.62	43.1 mg/dl	-0.11	0.39
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Biosystems	215	9.38	9.85	0.92	9.1 mg/dl	-0.30	0.13
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Biosystems	234	2.58	18.61	0.48	2.2 mg/dl	-0.79	0.06
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	237	5.33	8.82	0.47	5.4 g/dl	0.15	0.06
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	241	3.21	9.83	0.32	3 g/dl	-0.67	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Biosystems	225	8.31	9.48	0.79	8.05 mg/dl	-0.33	0.11
8	CHOLESTROL II	CHOD-PAP II	Biosystems	259	112.50	9.93	11.17	100 mg/dl	-1.12	1.39
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Biosystems	229	156.47	10.68	16.72	154 mg/dl	-0.15	2.21

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****