

VIEW LAB MONTHLY SUMMARY

Lab Name	PREM CLINICAL LAB	Details About Robust Analysis								
Lab No	13171	Detail About Monthly Summary								
Month	December	Detail about SDI								
Year	2021									
Constituent Group	Chemistry II									
Click on the analyte to view Graphical Data										
<div style="display: flex; justify-content: space-around;"> All Analyser Result Print </div>										
Date of Result Entered : 20/12/2021										
Date of Report Published : 03/01/2022										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Biosystems	283	216.52	9.32	20.17	221 mg/dl	0.22	2.40
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	762	105.93	13.00	13.77	113 mg/dl	0.51	1.00
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Biosystems	214	2.33	12.04	0.28	2.4 mg/dl	0.25	0.04
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Biosystems	217	3.15	13.35	0.42	2.9 mg/dl	-0.60	0.06
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	243	5.15	9.29	0.48	5 g/dl	-0.31	0.06
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	224	3.15	9.41	0.30	2.9 g/dl	-0.84	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Biosystems	210	4.26	17.03	0.72	4.13 mg/dl	-0.18	0.10
8	CHOLESTROL II	CHOD-PAP II	Biosystems	255	107.48	10.36	11.14	103 mg/dl	-0.40	1.40
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Biosystems	215	94.81	12.03	11.41	88 mg/dl	-0.60	1.56
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

