

VIEW LAB MONTHLY SUMMARY

Lab Name	PREM CLINICAL LAB	Details About Robust Analysis									
Lab No	13171	Detail About Monthly Summary									
Month	October	Detail about SDI									
Year	2021										
Constituent Group	Chemistry II										
		Click on the analyte to view Graphical Data				All Analyser Result		Print			
		Date of Result Entered : 20/10/2021									
		Date of Report Published : 01/11/2021									
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U	
1	GLUCOSE II	GOD-POD II	Biosystems	284	106.90	9.72	10.39	103 mg/dl	-0.38	1.23	
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	887	29.20	15.70	4.58	26.6 mg/dl	-0.57	0.31	
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Biosystems	229	1.39	12.11	0.17	1.5 mg/dl	0.65	0.02	
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Biosystems	235	7.17	11.82	0.85	6.8 mg/dl	-0.44	0.11	
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	241	5.18	8.98	0.46	5.3 g/dl	0.26	0.06	
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	233	3.15	9.78	0.31	2.9 g/dl	-0.81	0.04	
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Biosystems	241	6.08	12.26	0.75	6.23 mg/dl	0.20	0.10	
8	CHOLESTROL II	CHOD-PAP II	Biosystems	262	104.51	11.20	11.71	93 mg/dl	-0.98	1.45	
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Biosystems	221	106.95	12.81	13.71	106 mg/dl	-0.07	1.84	
		SDI Range		Interpretation							
		Within -1.00 to +1.00		Excellent.							
		Within ±1.01 to ±2.00		Good.							
		Within ±2.01 to ±2.99		Accept with caution. Warning Signal.							
		Beyond ±3.0		Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out