

LAB MONTHLY SUMMARY



Lab Name **PREM CLINICAL LAB** Lab No **13171**
 Month **June** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 20/06/2021

Date of Report Published : 03/07/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Biosystems	269	197.18	9.70	19.13	186 mg/dl	-0.58	2.33
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	871	74.34	15.22	11.31	78.5 mg/dl	0.37	0.77
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Biosystems	214	4.50	10.03	0.45	4.9 mg/dl	0.89	0.06
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Biosystems	228	4.04	15.18	0.61	4 mg/dl	-0.07	0.08
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	231	5.40	10.46	0.56	5.7 g/dl	0.53	0.07
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	231	3.27	10.33	0.34	2.9 g/dl	-1.09	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Biosystems	230	4.96	18.48	0.92	6 mg/dl	1.14	0.12
8	CHOLESTROL II	CHOD-PAP II	Biosystems	254	107.69	11.14	12.00	93 mg/dl	-1.22	1.51
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Biosystems	225	139.96	13.13	18.37	140 mg/dl	0.00	2.45

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****