

Department of Biochemistry

Internal Quality Control (IQC) & CV% (Biochemistry)

S.NO	Date	Nature of Problem	Root cause analysis and corrective and preventive actions	Date of Closure	Signature
1	03/01/2022 (Date of Result)	Value of CV% of Glucose was high	We have done the machine maintenance checked for all parameter's, Distilled water has been changed, Probe has been checked and changed, again all parameter has been recalibrated by the service engineer. Temperature record were checked. Then value reports came normal.	22/02/2022	Janta DM

IMPULSE
DIAGNOSTICS

