



ate Limited

HIN/MED/2021-2022/100410

05/10/ 2021

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **YumizenH550** bearing serial number: **910YAXH02141** installed at **PATHKIND LABS ,VARANASI** as calibrated on 10st SEPT 2021.

Calibrator : ABX MINOCAL
Lot No. : CX462.
Expiry Date : 5th OCT 2021

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on **9th SEPT 2022.**



Shrish Dixit

(Head- Products & Marketing)
For **Horiba India Pvt. Ltd.**