

## LAB MONTHLY SUMMARY



Lab Name **BHARATH COMPUTERISED  
CLINICAL LABORATORY** Lab No **11269**  
Month **June** Year **2021**  
Constituent Group **Chemistry II**



Date of Result Entered : 19/06/2021

Date of Report Published : 03/07/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	940	193.80	10.52	20.38	172.8 mg/dl	-1.03	1.33
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation)	871	74.34	15.22	11.31	75.8 mg/dl	0.13	0.77
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	675	4.42	11.68	0.52	4.9 mg/dl	0.93	0.04
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Any Analyser (Automation / Semi Automation)	800	3.86	16.46	0.64	3.9 mg/dl	0.06	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	729	5.51	11.11	0.61	6 g/dl	0.80	0.05
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	703	3.28	10.79	0.35	4 g/dl	2.03	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	756	5.04	22.70	1.14	7.5 mg/dl	2.16	0.08
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	906	109.30	12.80	13.99	119.7 mg/dl	0.74	0.93
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	774	139.56	14.86	20.74	135.1 mg/dl	-0.22	1.49

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102