

## LAB MONTHLY SUMMARY



Lab Name **BHARATH COMPUTERISED  
CLINICAL LABORATORY** Lab No **11269**  
Month **December** Year **2021**  
Constituent Group **Chemistry II**



Date of Result Entered : 19/12/2021

Date of Report Published : 03/01/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	969	215.80	9.25	19.96	190.2 mg/dl	-1.28	1.28
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation)	762	105.93	13.00	13.77	94.4 mg/dl	-0.84	1.00
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	725	2.24	11.51	0.26	2.5 mg/dl	1.01	0.02
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Any Analyser (Automation / Semi Automation)	789	3.00	13.68	0.41	3.2 mg/dl	0.49	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	738	5.21	9.17	0.48	5.6 g/dl	0.82	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	728	3.19	10.00	0.32	3.4 g/dl	0.66	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	776	4.39	19.07	0.84	3.7 mg/dl	-0.82	0.06
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	975	108.50	11.43	12.40	106.9 mg/dl	-0.13	0.79
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	803	96.54	15.20	14.68	108.3 mg/dl	0.80	1.04

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)

Contact Number: 0416-2283102