

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN: 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE



Customer Name & Address:	M/S. BHARATH Computerised Clinical Lab, #26, Gomathi Complex, Dr.Radhakrishnan Rd, Teachers Colony, Erode - 638 011.	Certificate No:	AMS/LAB/2889540
		Issuing Date:	19/07/2021
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Centrifuge	Id. No:	BCL/LAB/CEN/001
Make:	Remi	Department:	Laboratory
Model:	C852	Calibrated On:	17/07/2021
Sl. No:	ZCFN 16287	Calibration Due Suggested:	16/07/2022
Type:	Manual	DUC Condition on Receipt:	OK
Range & Least Count:	0 - 3500 / LC - 1		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Digital Tachometer	Mextech / DT-2234C	PC/01/21/ATL/0339-02	22/01/2022
2	Thermo Hygrometer	Beetech/088	PC/01/21/ATL/0194-01	22/01/2022

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL.No.	DUC Reading (RPM)	Standard Reading (RPM)	Correction (RPM)
1	Position 1	1116	0
2	Position 2	1460	0
3	Position 3	2189	0
4	Position 4	2568	0
5	Position 5	2997	0

Allowable Tolerance: Set by Customer.

Note:

1. Calibration Points were selected as per customer request.
2. Each reading is taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.



Calibrated By:
(Calibration Engineer)

B. S. S.



Approved by:
(Technical Manager)

B. S. S.