

**Analyte Methodwise Result Report**

Lab : 455  
 Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
 Report Date : 19/03/2021

Cycle No : 210103  
 Sample No : 1  
 Sample Date : 03-03-2021


**Haematology - CBC WB, Eosinophils Differential., %**

	N	Mean	SD	CV%	Excluded
<span style="color: purple;">■</span> All Methods	4	5.75	0.50	8.70	--
<span style="color: teal;">■</span> Impedance	4	5.75	0.50	8.70	--
<span style="color: darkred;">■</span> ERBA H 360	1	--	--	--	--

<b>Your Result</b>	6.00	<b>SDI / z-Score</b>	0.00
<b>Assigned Value</b>	6.00	<b>Average SDI</b>	0.00
<b>SDPA</b>	0.500	<b>Uncertainty of AV</b>	0.31
<b>Standard Unit Result Val</b>	6.00	<b>%DEV</b>	4.35

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Remarks: Interpretation of your reports

- If your SDI lies between +/- 2.0 SD then your results are well within limits.
- If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- If your SDI is +/- 3.0 SD or more then it is an action alert flag.

Analyte Methodwise Result Report

Lab : 455  
 Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
 Report Date : 19/03/2021

Cycle No : 210103  
 Sample No : 1  
 Sample Date : 03-03-2021

Haematology - CBC WB, Eosinophils Differential.					
	N	Mean	SD	CV%	Excluded
All Analyzer	4	5.75	0.50	8.70	--
<b>Your Analyzer :</b> ERBA H 360	1	--	--	--	--
ERBA XN 550	1	--	--	--	--
ERBA H 560	1	--	--	--	--
Sysmex XS 800 I	1	--	--	--	--

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
- Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
- All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.

**Analyte Methodwise Result Report**

Lab : 455  
Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
Report Date : 19/03/2021

Cycle No : 210103  
Sample No : 1  
Sample Date : 03-03-2021


**Haematology - CBC WB, Erythrocyte (RBC) Count., mill/cu.mm**

	N	Mean	SD	CV%	Excluded
<span style="color: purple;">■</span> All Methods	4	2.65	0.29	10.87	1
<span style="color: teal;">■</span> Impedance	4	2.65	0.29	10.87	1
<span style="color: red;">■</span> ERBA H 360	1	--	--	--	--

<b>Your Result</b>	2.73	<b>SDI / z-Score</b>	0.48
<b>Assigned Value</b>	2.61	<b>Average SDI</b>	0.14
<b>SDPA</b>	0.249	<b>Uncertainty of AV</b>	0.16
<b>Standard Unit Result Val</b>	2.73	<b>%DEV</b>	3.02

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.

**Analyte Methodwise Result Report**

**Lab :** 455  
**Lab Name :** MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
**Report Date :** 19/03/2021

**Cycle No :** 210103  
**Sample No :** 1  
**Sample Date :** 03-03-2021

<b>Haematology - CBC WB, Erythrocyte (RBC) Count.</b>					
	<b>N</b>	<b>Mean</b>	<b>SD</b>	<b>CV%</b>	<b>Excluded</b>
All Analyzer	4	2.65	0.29	10.87	1
<b>Your Analyzer :</b> ERBA H 360	1	--	--	--	--
ERBA XN 550	1	--	--	--	--
ERBA H 560	1	--	--	--	1
Sysmex XS 800 I	1	--	--	--	--

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
- Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
- All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.

**Analyte Methodwise Result Report**

Lab : 455  
Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
Report Date : 19/03/2021

Cycle No : 210103  
Sample No : 1  
Sample Date : 03-03-2021


**Haematology - CBC WB, Haemoglobin., g/dL**

	N	Mean	SD	CV%	Excluded
<span style="color: purple;">■</span> All Methods	4	6.93	0.15	2.21	1
<span style="color: teal;">■</span> Impedance	4	6.93	0.15	2.21	1
<span style="color: red;">■</span> ERBA H 360	1	--	--	--	--

<b>Your Result</b>	6.90	<b>SDI / z-Score</b>	0.17
<b>Assigned Value</b>	6.85	<b>Average SDI</b>	0.09
<b>SDPA</b>	0.294	<b>Uncertainty of AV</b>	0.18
<b>Standard Unit Result Val</b>	6.90	<b>%DEV</b>	-0.48

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.

Analyte Methodwise Result Report

Lab : 455  
 Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
 Report Date : 19/03/2021

Cycle No : 210103  
 Sample No : 1  
 Sample Date : 03-03-2021

Haematology - CBC WB, Haemoglobin.					
	N	Mean	SD	CV%	Excluded
All Analyzer	4	6.93	0.15	2.21	1
<b>Your Analyzer : ERBA H 360</b>	1	--	--	--	--
ERBA XN 550	1	--	--	--	--
ERBA H 560	1	--	--	--	1
Sysmex XS 800 I	1	--	--	--	--

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
- Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
- All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.

**Analyte Methodwise Result Report**

Lab : 455  
Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
Report Date : 19/03/2021

Cycle No : 210103  
Sample No : 1  
Sample Date : 03-03-2021


**Haematology - CBC WB, Lymphocytes Differential., %**

	N	Mean	SD	CV%	Excluded
<span style="color: purple;">■</span> All Methods	4	30.67	5.13	16.74	1
<span style="color: teal;">■</span> Impedance	4	30.67	5.13	16.74	1
<span style="color: red;">■</span> ERBA H 360	1	--	--	--	--

<b>Your Result</b>	32.00	<b>SDI / z-Score</b>	-0.21
<b>Assigned Value</b>	33.50	<b>Average SDI</b>	-0.08
<b>SDPA</b>	7.047	<b>Uncertainty of AV</b>	4.40
<b>Standard Unit Result Val</b>	32.00	<b>%DEV</b>	4.35

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Remarks: Interpretation of your reports

- If your SDI lies between +/- 2.0 SD then your results are well within limits.
- If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- If your SDI is +/- 3.0 SD or more then it is an action alert flag.

Analyte Methodwise Result Report

Lab : 455  
 Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
 Report Date : 19/03/2021

Cycle No : 210103  
 Sample No : 1  
 Sample Date : 03-03-2021

Haematology - CBC WB, Lymphocytes Differential.						
		N	Mean	SD	CV%	Excluded
	All Analyzer	4	30.67	5.13	16.74	1
<b>Your Analyzer :</b>	<b>ERBA H 360</b>	1	--	--	--	--
	ERBA XN 550	1	--	--	--	--
	ERBA H 560	1	--	--	--	1
	Sysmex XS 800 I	1	--	--	--	--

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
- Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
- All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.



**Analyte Methodwise Result Report**

Lab : 455  
Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
Report Date : 19/03/2021

Cycle No : 210103  
Sample No : 1  
Sample Date : 03-03-2021


**Haematology - CBC WB, MCH (Mean Corpuscular Hb)., pg**

	N	Mean	SD	CV%	Excluded
<span style="color: purple;">■</span> All Methods	4	26.40	2.46	9.31	1
<span style="color: teal;">■</span> Calculated	4	26.40	2.46	9.31	1
<span style="color: red;">■</span> ERBA H 360	1	--	--	--	--

<b>Your Result</b>	25.40	<b>SDI / z-Score</b>	-0.05
<b>Assigned Value</b>	25.50	<b>Average SDI</b>	0.19
<b>SDPA</b>	2.046	<b>Uncertainty of AV</b>	1.28
<b>Standard Unit Result Val</b>	25.40	<b>%DEV</b>	-3.79

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Remarks: Interpretation of your reports

- If your SDI lies between +/- 2.0 SD then your results are well within limits.
- If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- If your SDI is +/- 3.0 SD or more then it is an action alert flag.

**Analyte Methodwise Result Report**

**Lab :** 455  
**Lab Name :** MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
**Report Date :** 19/03/2021

**Cycle No :** 210103  
**Sample No :** 1  
**Sample Date :** 03-03-2021

<b>Haematology - CBC WB, MCH (Mean Corpuscular Hb).</b>					
	<b>N</b>	<b>Mean</b>	<b>SD</b>	<b>CV%</b>	<b>Excluded</b>
All Analyzer	4	26.40	2.46	9.31	1
<b>Your Analyzer :</b> ERBA H 360	1	--	--	--	--
ERBA XN 550	1	--	--	--	--
ERBA H 560	1	--	--	--	1
Sysmex XS 800 I	1	--	--	--	--

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
- Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
- All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.

**Analyte Methodwise Result Report**

Lab : 455  
Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
Report Date : 19/03/2021

Cycle No : 210103  
Sample No : 1  
Sample Date : 03-03-2021


**Haematology - CBC WB, MCHC (Mean Corpuscular Hb Concentration), g/dL**

	N	Mean	SD	CV%	Excluded
<span style="color: purple;">■</span> All Methods	4	30.40	3.12	10.26	1
<span style="color: teal;">■</span> Calculated	4	30.40	3.12	10.26	1
<span style="color: red;">■</span> ERBA H 360	1	--	--	--	--

<b>Your Result</b>	28.70	<b>SDI / z-Score</b>	0.03
<b>Assigned Value</b>	28.60	<b>Average SDI</b>	0.90
<b>SDPA</b>	2.883	<b>Uncertainty of AV</b>	1.80
<b>Standard Unit Result Val</b>	28.70	<b>%DEV</b>	-5.59

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.

Analyte Methodwise Result Report

Lab : 455  
 Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
 Report Date : 19/03/2021

Cycle No : 210103  
 Sample No : 1  
 Sample Date : 03-03-2021

Haematology - CBC WB, MCHC (Mean Corpuscular Hb Concentration).					
	N	Mean	SD	CV%	Excluded
All Analyzer	4	30.40	3.12	10.26	1
<b>Your Analyzer :</b> ERBA H 360	1	--	--	--	--
ERBA XN 550	1	--	--	--	--
ERBA H 560	1	--	--	--	1
Sysmex XS 800 I	1	--	--	--	--

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
- Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
- All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.

**Analyte Methodwise Result Report**

Lab : 455  
Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
Report Date : 19/03/2021

Cycle No : 210103  
Sample No : 1  
Sample Date : 03-03-2021


**Haematology - CBC WB, MCV (Mean Corpuscular Volume), fL**

	N	Mean	SD	CV%	Excluded
<span style="color: purple;">■</span> All Methods	4	86.80	1.40	1.61	1
<span style="color: teal;">■</span> Calculated	4	86.80	1.40	1.61	1
<span style="color: red;">■</span> ERBA H 360	1	--	--	--	--

<b>Your Result</b>	88.40	<b>SDI / z-Score</b>	0.36
<b>Assigned Value</b>	87.30	<b>Average SDI</b>	-0.09
<b>SDPA</b>	3.071	<b>Uncertainty of AV</b>	1.92
<b>Standard Unit Result Val</b>	88.40	<b>%DEV</b>	1.84

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.

**Analyte Methodwise Result Report**

**Lab :** 455  
**Lab Name :** MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
**Report Date :** 19/03/2021

**Cycle No :** 210103  
**Sample No :** 1  
**Sample Date :** 03-03-2021

<b>Haematology - CBC WB, MCV (Mean Corpuscular Volume).</b>					
	<b>N</b>	<b>Mean</b>	<b>SD</b>	<b>CV%</b>	<b>Excluded</b>
All Analyzer	4	86.80	1.40	1.61	1
<b>Your Analyzer :</b> ERBA H 360	1	--	--	--	--
ERBA XN 550	1	--	--	--	--
ERBA H 560	1	--	--	--	1
Sysmex XS 800 I	1	--	--	--	--

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
- Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
- All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.

**Analyte Methodwise Result Report**

Lab : 455  
Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
Report Date : 19/03/2021

Cycle No : 210103  
Sample No : 1  
Sample Date : 03-03-2021


**Haematology - CBC WB, Neutrophils Differential., %**

	N	Mean	SD	CV%	Excluded
<span style="color: purple;">■</span> All Methods	4	59.67	3.79	6.35	1
<span style="color: teal;">■</span> Impedance	4	59.67	3.79	6.35	1
<span style="color: red;">■</span> ERBA H 360	1	--	--	--	--

<b>Your Result</b>	58.00	<b>SDI / z-Score</b>	0.07
<b>Assigned Value</b>	57.50	<b>Average SDI</b>	0.07
<b>SDPA</b>	7.047	<b>Uncertainty of AV</b>	4.40
<b>Standard Unit Result Val</b>	58.00	<b>%DEV</b>	-2.79

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.

Analyte Methodwise Result Report

Lab : 455  
 Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
 Report Date : 19/03/2021

Cycle No : 210103  
 Sample No : 1  
 Sample Date : 03-03-2021

Haematology - CBC WB, Neutrophils Differential.					
	N	Mean	SD	CV%	Excluded
All Analyzer	4	59.67	3.79	6.35	1
<b>Your Analyzer :</b> ERBA H 360	1	--	--	--	--
ERBA XN 550	1	--	--	--	--
ERBA H 560	1	--	--	--	1
Sysmex XS 800 I	1	--	--	--	--

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
- Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
- All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.



**Analyte Methodwise Result Report**

Lab : 455  
Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
Report Date : 19/03/2021

Cycle No : 210103  
Sample No : 1  
Sample Date : 03-03-2021


**Haematology - CBC WB, Platelet Count., 1000/  $\mu$ L**

	N	Mean	SD	CV%	Excluded
<span style="color: purple;">■</span> All Methods	4	290.67	30.17	10.38	1
<span style="color: teal;">■</span> Impedance	4	290.67	30.17	10.38	1
<span style="color: red;">■</span> ERBA H 360	1	--	--	--	--

<b>Your Result</b>	305.00	<b>SDI / z-Score</b>	-0.06
<b>Assigned Value</b>	308.00	<b>Average SDI</b>	0.21
<b>SDPA</b>	47.546	<b>Uncertainty of AV</b>	29.72
<b>Standard Unit Result Val</b>	305.00	<b>%DEV</b>	4.93

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.

Analyte Methodwise Result Report

Lab : 455  
 Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
 Report Date : 19/03/2021

Cycle No : 210103  
 Sample No : 1  
 Sample Date : 03-03-2021

Haematology - CBC WB, Platelet Count.						
		N	Mean	SD	CV%	Excluded
	All Analyzer	4	290.67	30.17	10.38	1
<b>Your Analyzer :</b>	<b>ERBA H 360</b>	1	--	--	--	--
	ERBA XN 550	1	--	--	--	--
	ERBA H 560	1	--	--	--	1
	Sysmex XS 800 I	1	--	--	--	--

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
- Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
- All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.

**Analyte Methodwise Result Report**

Lab : 455  
Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
Report Date : 19/03/2021

Cycle No : 210103  
Sample No : 1  
Sample Date : 03-03-2021


**Haematology - CBC WB, Total Leucocytes (WBC) count., /c.mm**

	N	Mean	SD	CV%	Excluded
<span style="color: purple;">■</span> All Methods	4	3,375.00	516.19	15.29	2
<span style="color: teal;">■</span> Impedance	4	3,375.00	516.19	15.29	2
<span style="color: darkred;">■</span> ERBA H 360	1	--	--	--	--

<b>Your Result</b>	3,010.00	<b>SDI / z-Score</b>	-0.23
<b>Assigned Value</b>	3,100.00	<b>Average SDI</b>	-0.23
<b>SDPA</b>	398.037	<b>Uncertainty of AV</b>	287.26
<b>Standard Unit Result Val</b>	3010.00	<b>%DEV</b>	-10.81

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.

**Analyte Methodwise Result Report**

**Lab :** 455  
**Lab Name :** MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
**Report Date :** 19/03/2021

**Cycle No :** 210103  
**Sample No :** 1  
**Sample Date :** 03-03-2021


<b>Haematology - CBC WB, Total Leucocytes (WBC) count.</b>					
	<b>N</b>	<b>Mean</b>	<b>SD</b>	<b>CV%</b>	<b>Excluded</b>
All Analyzer	4	3,375.00	516.19	15.29	2
<b>Your Analyzer :</b> ERBA H 360	1	--	--	--	--
ERBA XN 550	1	--	--	--	1
ERBA H 560	1	--	--	--	1
Sysmex XS 800 I	1	--	--	--	--

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
- Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
- All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SD then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SD then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SD or more then it is an action alert flag.



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
 Unit No. 409-416,  
 Commercial Building - 1A,  
 Kohinoor Mall, Kirol Road, Kurla (W),  
 Mumbai - 400070