

SER/F/001/00

INSTALLATION REPORT



CUSTOMER DETAILS : (* TO BE FILLED IN CAPITAL LETTERS)

Customer Name : SP Hitech path lab .
 Address : Trimurti plaza .
Talegaon Dhabade . Tal. Naval .
 Pin code : 410506 . City : Pune . State : Maharashtra .
 Contact Person : Dr. Sachin Tonape .
 Mobile/ Tel : 9921248229 Email : _____

INSTRUMENT DETAILS :

Model : AGD300 . Date of Installation : 6/9/21 .
 Instrument Serial No. : 17410707210744 . Warranty Expiry Date : 5/9/22 .
 Software Version : _____ Invoice No. : _____
 Invoiced from AGD/Distributor (If through distributor, please mention the name)
 Distributor Name : _____ Place : Pune

PHYSICAL INSPECTION & PRE-INSTALLATION CHECKS :

Instrument received in good condition : Yes No Vibration free Platform : Yes No
 CVT/Online UPS available : Yes No Dust Free : Yes No
 Is room Air-conditioned : Yes No Away from Sunlight : Yes No
 Mains AC supply : L & N : 238 VAC N & E : 2 VAC L & E : 236 VAC

ENGINEER REMARKS :

Done Installation

TO BE FILLED BY THE CUSTOMER :

We agree that the above mentioned instrument has been received in good condition and installed to my satisfaction by the Engineer. I/We have been given operational and maintenance training by the Engineer to our satisfaction.

Customer's Comments, If any _____

Anam
 Signature of Customer :
 Seal & Date



Sachin Tonape
 Name and Signature of Engineer

AGD Biomedicals (P) Ltd. P : +91-22-28257999 ISO : 9001 : 2015
 Mehta Trade Center, +91-22-28257999 ISO : 13485 : 2016
 Sir M. V. Road, Andheri East E : sales@agdbio.com
 Mumbai - 400 099 INDIA W : www.agdbio.com