



### CALIBRATION CERTIFICATE

Certificate No.	MES/21-22/GOWOCHMA/REF/J03	Date of Issue	: 06/08/2021
		Calibration Done On	: 03/08/2021
		Calibration Due On	: 02/08/2022
Customer Name	GOVERNMENT WOMEN & CHILDREN'S HOSPITAL, MATTANCHERRY, ERNAKULAM.		

#### DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Refrigerator	LG	-	ICTC/134/KSACS /01	ICTC

#### DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Thermometer	Metravi / DTM 900	10055593	ACS/20-21/I02669-01	23/09/2021
2	Digital Thermo Hygrometer	Equinox & EQ - 308 CTH	20191602	ACS/JUN/21/I - 32 -01	17/06/2022

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27 °C	Relative Humidity	55 %
--------------------------	-------------	-------	-------------------	------

#### PERFORMANCE DATA

	Measured Reading	Range
Body	6.2	2 to 8 °C
Freezer	-13.5	-20 to -10 °C

Calibration Status	Pass	✓	Fail
--------------------	------	---	------

Calibrated By

Sajin (Service Engineer)

Approved By



Binoj Thomas (Quality Manager)

NB: Results reported are valid at the time of and under the stated conditions of measurements. This report shall not be reproduced without written approval by Medical Engineering & Services.