

ST-200 aQua CALIBRATION CERTIFICATE

CUSTOMERS NAME: Likhithas Diagnostics and Speciality Lab			
ADDRESS: H. No. 1-8-65/A, Nort Hyderabad -500062.	h Kamalanagar, Above Andhra Bank, ECIL,		
CONTACT NO: 8639830362			
SERIAL NO OF THE INSTRUMENT:	ST-06221(F)		
CALIBRATION DATE: 08.07.2021	VALID TILL: 07.07.2022		

CALIBRATION DONE BY: Mr. KIRAN DATE

Calibration	Slopes:			
S.No.	Na+	K+	iCa	Cl ⁻
1	64.69	63.56	34.78	50.89
2	64.80	63.54	35.00	51.02
3	64.88	63.62	35.08	51.06
Mean	64.79	63.57	34.95	50.99
Ranges	35 - 100mV/dec	35-100 mV/dec	15 - 50 mV/dec	30 - 100 mV/dec
Status	O.K	O.K	O.K	O.K



Quality Control L1 :	Lot:- UQCA2103A – L	1		Exp: 28.02.2023
S.No.	Na+	K +	iCa	Cl¯
1	123.4	1.61	1.86	84.6
2	122.7	1.65	1.85	84.9
3	124.9	1.73	1.84	88.7
Mean	123.67	1.66	1.85	86.07
Ranges	110 - 128 mmol/L	1.46 - 2.46 mmol/L	1.34 - 2.34 mmol/L	72 – 90 mmol/L
Status	O.K	O.K	O.K	O.K
Quality Control 2:	Lot:- UQCA2103A -L	2		Exp: 28.02.2023
S.No.	Na+	K +	iCa	Cl
1				
1	143.2	4.09	1.32	104.6
3	143.1	4.08	1.32	104.7
	143.1	4.08	1.32	104.7
Mean	143.13	4.08	1.32	104.67
Ranges	134 – 148 mmol/L	3.4 - 4.4 mmol/L	0.70 - 1.70 mmol/L	94 - 108 mmol/L
Status	O.K	O.K	O.K	O.K
Quality Control 3:	Lot:- UQCA2103A -L	2		Exp: 28.02.2023
S.No.	Na+	K +	iCa	Cl⁻
1	163.7	6.62	0.87	124.9
2	164	6.64	0.87	125
3	164.4	6.67	0.86	125.2
Mean	164.03	6.64	0.87	125.03
Ranges	150 - 168 mmol/L	5.40 - 6.80 mol/L	0.21 - 1.21 mmol/L	111 – 129 mmol/L
Status	O.K	O.K	O.K	O.K



Signature of Engineer

Signature of Doctor

Date: 08.07.2021 Date: 08.07.2021 ANALYZE TYPE: QUALITY CONTROL Jul-08-21 10:46:07

SENSA	CORE
LEVEL	1

ANALYZE QC

Na = 123.4 mmol/L K = 1.61 mmol/L iCa=1.86 mmol/L Cl = 84.6 mmol/L

mmol/L HIGH

HIGH

HIGH

Jul-08-21 10:59:34

SENSACORE LEVEL 1

ANALYZE QC

Na = 122.7 mmol/L K = 1.65 mmol/L iCa=1.85 mmol/L

C1 :84.9 mmol/L

Jul-08-21 11:01:25

SENSACORE LEVEL 1

ANALYZE QC

Na = 124.9 mmol/L K = 1.73 mmol/L iCa=1.84 mmol/L Cl = 88.7 mmol/L

Jul-08-21 11:03:14

SENSACORE LEVEL 2

ANALYZE QC

Na = 143.2 mmol/L HIGH K = 4.09 mmol/L HIGH iCa=1.32 mmol/L HIGH CI = 104.6 mmol/L HIGH

Jul-08-21 11:05:09

SENSACORE LEVEL 2

ANALYZE QC

Na = 143.1 mmol/L HIGH K = 4.08 mmol/L HIGH iCa=1.32 mmol/L HIGH CI = 104.7 mmol/L HIGH

Jul-08-21 11:06:54

SENSACORE LEVEL 2

ANALYZE QC

Na = 143.1 mmol/L HIGH K = 4.08 mmol/L HIGH iCa=1.32 mmol/L HIGH CI = 104.7 mmol/L HIGH

とうというないというとうないとうというなりというというというというと

Jul-08-21 11:09:13

SENSACORE LEVEL 3

ANALYZE QC

Na :163.7 mmol/L HIGH K :6.62 mmol/L HIGH iCa:0.87 mmol/L HIGH Cl :124.9 mmol/L HIGH

Jul-08-21 11:11:05

SENSACORE LEVEL 3

ANALYZE QC

Na :164.0 mmol/L HIGH K :6.64 mmol/L HIGH iCa:0.87 mmol/L HIGH CI:125.0 mmol/L HIGH

Jul-08-21 11:14:06

SENSACORE LEVEL 3

ANALYZE QC

Na :164.4 mmol/L K :6.67 mmol/L iCa:0.86 mmol/L Cl:125.2 mmol/L

Jul-08-21 11:16:07

and the second s

1202/.

HIGH

HIGH

HIGH

HIGH «



INSTALLATION QUALIFICATION – IQ OPERATIONAL QUALIFICATION – OQ

&

PERFORMANCE QUALIFICATION - PQ



ST-200 aQua Electrolyte Analyzer

INSTALLATION QUALIFICATON

FOR

Likhitha's Diagnostics & Speciality Lab

Kapra sainikpuri, Opp.Srikara Hospital, Bedside Andhra Bank, ECIL, North Kamala Nagar, Hyderabad, Telangana 500062



CERTIFICATION OF OPERATION

INSTRUMENT NAME : ST-200aQua

SERIAL NUMBER : STAQW-06221

CUSTOMER NAME : Likhitha's Diagnostic Lab.

ADDRESS : ECIL, Hyderabad

THE UNDERSIGNED PERFORMERS CERTIFY THAT THE OPERATIONAL QUALIFICATION PROTOCOL HAS BEEN SUCCESSFULLY COMPLETED FOR THE INSTRUMENT STATED ABOVE.

ENGINEER

Signature :

Name : MR. KIRAN DATE

Designation : NATIONAL SERVICE MANAGER

Company : SENSACORE MEDICAL INSTRUMENTATION Pvt. Ltd.

POS Bala

QUALITY MANAGER

Signature

Name

Designation : QA/QC TEAM LEADER

Company : SENSACORE MEDICAL INSTRUMENTATION Pvt. Ltd.

MR. POSI BABU V



OPERATIONAL QUALIFICATION OF THE INSTRUMENT:

DESCRIPTION	STATUS	REMARKS
POWER SUPPLY	SELF CHECK	PASSED
ELECTRONICS	SELF CHECK	PASSED
REAGENT PACK	SELF CHECK	PASSED
HARDWARE	SELF CHECK	PASSED
FLOW PATH	SELF CHECK	PASSED





ST-200 aQua Electrolyte Analyzer

OPERATIONAL QUALIFICATION

FOR

Likhitha's Diagnostics & Speciality Lab

Kapra sainikpuri, Opp.Srikara Hospital, Bedside Andhra Bank, ECIL, North Kamala Nagar, Hyderabad, Telangana 500062



CERTIFICATION OF INSTALLATION

INSTRUMENT NAME : ST-200 aQua

SERIAL NUMBER : STAQW-6221

CUSTOMER NAME: Likhitha's Diagnostic Lab.

ADDRESS : ECIL, Hyderabad

THE UNDERSIGNED PERFORMERS CERTIFY THAT THE INSTALLATION QUALIFICATION PROTOCOL HAS BEEN SUCCESSFULLY COMPLETED FOR THE INSTRUMENT STATED ABOVE.

ENGINEER

Signature

Name

-

Designation : NATIONAL SERVICE MANAGER

MR. KIRAN DATE

Company : SENSACORE MEDICAL INSTRUMENTATION Pvt. Ltd.

po So Balan

QUALITY MANAGER

Signature :

Name : MR. POSI BABU V

Designation : QA/QC TEAM LEADER

Company : SENSACORE MEDICAL INSTRUMENTATION Pvt. Ltd.



INSTALLATION REPORT

CUSTOMER NAME : Likhitha's Diagnostic

PLACE : ECIL , Hyderabad

DEPARTMENT : Pathology Laboratory

INSTRUMENT NAME : ST 200 aQua

SERIAL NUMBER : STAQW-06221

INSTALLATION DATE : 09.07.2019

WARRANTY EXPIRY : 08.07.2022

THE ELECTROLYTE ANALYZER WAS INSTALLED ALONG WITH THE NECESSARY STANDARD ACCESSORIES, ALL BASIC REQUIREMENTS AND ADJUSTMENTS WERE CHECKED AND FOUND TO BE SATISFACTORY.THE PRELIMINARY STANDARDIZATION OF THE ELECTROLYTE ANALYZER AND THE TRAINING WERE PROVIDED TO THE COMPLETE SATISFACTION OF THE CUSTOMER. THE ELECTROLYTE ANALYZER IS FOUND TO BE WORKING SATISFACTORILY.





INSTRUMENT SETUP

S No	ITEM DESCRIPTION	STATUS
1	SOFTWARE VERSION	INSTALLED
2	REAGENT PACK	INSTALLED
3	QUALITY CONTROL	TESTED

OPERATION INSPECTION

- POWER SUPPLY CHECKED.
- NO UNUSUAL NOISE PRODUCED IN THE SYSTEM.





INSTALLATION QUALIFICATION

INSTRUMENT NAME: S

ST-200aQua

SERIAL NUMBER

STAQW-06221

CUSTOMER NAME

Likhitha's Diagnostic Lab.

ADDRESS

ECIL, Hyderabad

INSTALLATION REQUIREMENTS FOR THE INSTRUMENT:

REQUIREMENTS	ACCEPTABLE RANGE	OBSERVED RANGE
Input Voltage	220 <u>+</u> 10AC	217 <u>+</u> 5V AC
Line Frequency	50/60 HZ	50HZ
Ambient Temperature	10-35 degree	24 degree





ST-200aQua Electrolyte Analyzer

PERFORMANCE QUALIFICATION

FOR

Likhitha's Diagnostics & Speciality Lab

Kapra sainikpuri, Opp.Srikara Hospital, Bedside Andhra Bank, ECIL, North Kamala Nagar, Hyderabad, Telangana 500062



PERFORMANCE ANALYSIS

The performance of the system has been verified by the following process.

Ten replicates of the control Level 1 & Level 2 & Level 3 have been run and the statistical analysis of the data is performed. The results should be within control range with following precision limits.

THE UNDERSIGNED PERFORMERS CERTIFY THAT THE PERFORMANCE QUALIFICATION PROTOCOL HAS BEEN SUCCESSFULLY COMPLETED FOR THE INSTRUMENT STATED ABOVE.

A separate annexure has been attached along with this document

ENGINEER

Signature

Signature

Name : MR. KIRAN DATE

Designation : NATIONAL SERVICE MANAGER

Company : SENSACORE MEDICAL INSTRUMENTATION Pvt. Ltd.

pos Bala

QUALITY MANAGER

Name : MR. POSI BABU V

Designation : QA/QC TEAM LEADER

Company : SENSACORE MEDICAL INSTRUMENTATION Pvt. Ltd.

Likhithas Diagnostics and Speciality Lab, ECIL, Hyderabad

Electrolytes Precision Study

Equipment : Sensa Core ST200 aQua Equipment S.N. : STAQW - 06221

Date: 10.02.2022 Pooled Patient Serum Sample

S.N.	SODIUM	POTASSIUM	CHLORIDE
1	139.7	4.72	100.3
2	139.6	4.77	102.4
3	141.2	4.86	102.0
4	141.0	4.90	101.7
5	141.2	4.93	102.4
Mean	140.5	4.84	101.8
SD	0.82	0.09	0.87
%CV	0.58	1.83	0.85
MU	1.14	3.59	1.67
CLIA Specifications	± 4 mmol/L	± 0.3 mmol/L	± 5%

Processed By: Mr. Ch. Vamshi, Sr. Lab Technician

Verified By : Quality Manager

PATIENT ID:01111 01 SENSACORE ANALYZE SAMPLE Na =139.7 mmol/L K =4.72 mmol/L mmol/L mmol/L CI =100.3 mmol/L Feb-10-22 15:00:20 PATIENT ID:022222 SENSACORE ANALYZE SAMPLE Na = 139.6 mmol/L K =4.77 mmol/L iCa=1.13 mmol/L CI =102.4 mmol/L Feb-10-22 15:02:26 PATIENT ID:033333 SENSACORE ANALYZE SAMPLE Na :141 . 2 mmol/L K =4 -86 mmol/L iCa=1 - 15 mmol/L C =102.0 mmol/L Feb-10-22 15:04:07 PATIENT ID:044444 04 SENSACORE ANALYZE SAMPLE Na = 141 - O mmol/L K =4.90 mmol/L iCa=1.16 mmol/L CI =101.7 mmol/L Feb-10-22 15:05:48 PATIENT ID:055555 SENSACORE ANALYZE SAMPLE Na = 141 . 2 mmol/L K =4 - 93 mmol/L iCa=1 . 15 mmol/L 1 =102 - 4 mmol/L F-b-10-22 15:07:28

Halifill & Application of the second