



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JANUARY 2022



Lab Name **DR APARNA'S PATHOLOGY LABORATORY** Lab No **15380**  
 Constituent Group **Chemistry II** Date of Result Entered : **19/01/2022**  
 PT item **Lyophilized Serum** Date of Report Published : **07/02/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	Dry Chemistry	Fuji Dry Chemistry series	17	121.17	5.25	6.36	125 mg/dl	0.60	3.09
2	UREA II	Dry Chemistry	Fuji Dry Chemistry series	23	47.65	7.12	3.39	50.72 mg/dl	0.91	1.41
3	CREATININE II	Dry Chemistry	Fuji Dry Chemistry series	20	1.81	8.39	0.15	1.73 mg/dl	-0.53	0.07
4	T.BILIRUBIN II	Dry Chemistry	Fuji Dry Chemistry series	23	2.22	11.75	0.26	2.2 mg/dl	-0.08	0.11
5	T-PROTEIN II	Dry Chemistry	Fuji Dry Chemistry series	25	4.89	6.85	0.34	5.1 g/dl	0.63	0.13
6	ALBUMIN II	Dry Chemistry	Fuji Dry Chemistry series	25	3.08	6.00	0.18	3.2 g/dl	0.65	0.07
7	URIC ACID II	Dry Chemistry	Fuji Dry Chemistry series	29	6.98	6.16	0.43	7 mg/dl	0.05	0.16
8	CHOLESTROL II	Dry Chemistry	Fuji Dry Chemistry series	16	101.57	7.69	7.81	105 mg/dl	0.44	3.91
9	TRIGLYCERIDE II	Dry Chemistry	Fuji Dry Chemistry series	21	129.76	9.36	12.15	140 mg/dl	0.84	5.30

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

**DR APARNA'S PATHOLOGY LABORATORY**  
**SHOP 12, DHEERAJ GARDEN CHS, OPP ALN PARK, POONAM NAGAR, ANDHARI**  
**MUMBAI**  
**MAHARASHTRA400093**

*Pamela Christudoss*

**Coordinator Contact Details:**  
**Email:** [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
**Contact Number:** 0416-2283102

**Dr. Pamela Christudoss**  
**CMC EQAS Coordinator**  
**Christian Medical College, Vellore**

**Homogeneity and Stability of the sample is passed.**

**Data in CMC EQAS reports is confidential**

**CMC EQAS does not sub contract any components**

**\*\*\*\*\* End of Report \*\*\*\*\***