



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JANUARY 2022

Lab Name **DR APARNA'S PATHOLOGY LABORATORY** Lab No **15380**
 Constituent Group **Urine (Microalbumin & Protein)** Date of Result Entered : **20/01/2022**
 PT item **Lyophilized Urine** Date of Report Published : **07/02/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	UR. PROTEIN	Ur. Pyrogallol	Any Analyser (Automation / Semi Automation)	45	37.46	19.36	7.25	38.2 mg/dL	0.10	2.16
2	UR. MICROALBUMIN	Ur. Immuno Turbidimetric	Any Analyser (Automation / Semi Automation)	57	207.24	13.61	28.20	130.4 mg/L	-2.72	7.47

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

DR APARNA'S PATHOLOGY LABORATORY
SHOP 12, DHEERAJ GARDEN CHS, OPP ALN PARK, POONAM NAGAR, ANDHARI
MUMBAI
MAHARASHTRA400093

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

