

LAB MONTHLY SUMMARY



Lab Name **DAGAR DIAGNOSTIC CENTRE** Lab No **14266**
 Month **August** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 17/08/2021

Date of Report Published : 01/09/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	691	253.55	11.35	28.78	250.4 mg/dl	-0.11	2.19
2	UREA II	Urease UV / GLDH II	Transasia / Erba	580	43.59	12.70	5.53	44.7 mg/dl	0.20	0.46
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	472	1.96	12.60	0.25	2.02 mg/dl	0.24	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	618	3.97	11.54	0.46	4.03 mg/dl	0.13	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	619	5.52	9.17	0.51	5.93 g/dl	0.81	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	594	3.29	8.83	0.29	3.31 g/dl	0.07	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	643	3.54	22.19	0.78	3.5 mg/dl	-0.05	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	674	107.45	11.66	12.52	101 mg/dl	-0.52	0.96
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	629	120.29	15.63	18.81	107.4 mg/dl	-0.69	1.50

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****

LAB MONTHLY SUMMARY



Lab Name **DAGAR DIAGNOSTIC CENTRE** Lab No **14266**
 Month **September** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 20/09/2021

Date of Report Published : 05/10/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	770	399.21	11.44	45.67	439.3 mg/dl	0.88	3.29
2	UREA II	Urease UV / GLDH II	Transasia / Erba	584	90.64	11.10	10.06	95 mg/dl	0.43	0.83
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	483	3.37	11.05	0.37	3.63 mg/dl	0.70	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	670	2.10	15.07	0.32	2.22 mg/dl	0.38	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	597	5.39	8.88	0.48	5.65 g/dl	0.54	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	616	3.22	9.51	0.31	3.26 g/dl	0.13	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	675	8.55	14.14	1.21	8.1 mg/dl	-0.37	0.09
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	713	112.13	10.78	12.09	101 mg/dl	-0.92	0.91
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	658	217.36	11.08	24.09	204.2 mg/dl	-0.55	1.88

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****

LAB MONTHLY SUMMARY



Lab Name **DAGAR DIAGNOSTIC CENTRE** Lab No **14266**
 Month **October** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 20/10/2021

Date of Report Published : 01/11/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	803	105.70	10.78	11.39	105.8 mg/dl	0.01	0.80
2	UREA II	Urease UV / GLDH II	Transasia / Erba	614	27.49	15.59	4.28	25.9 mg/dl	-0.37	0.35
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	488	1.31	11.94	0.16	1.5 mg/dl	1.21	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	692	6.97	13.68	0.95	7.41 mg/dl	0.46	0.07
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	637	5.42	10.60	0.57	5.62 g/dl	0.35	0.05
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	636	3.21	9.77	0.31	3.22 g/dl	0.03	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	689	5.97	13.21	0.79	6.3 mg/dl	0.42	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	651	101.83	11.23	11.43	93 mg/dl	-0.77	0.90
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	570	92.74	16.83	15.61	84.4 mg/dl	-0.53	1.31

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****

LAB MONTHLY SUMMARY



Lab Name **DAGAR DIAGNOSTIC CENTRE** Lab No **14266**
 Month **November** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 19/11/2021

Date of Report Published : 03/12/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	775	144.88	10.50	15.21	141.3 mg/dl	-0.24	1.09
2	UREA II	Urease UV / GLDH II	Transasia / Erba	568	42.79	11.22	4.80	47 mg/dl	0.88	0.40
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	484	9.03	13.47	1.22	8.83 mg/dl	-0.16	0.11
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendressik II	Transasia / Erba	699	2.27	16.68	0.38	2.1 mg/dl	-0.45	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	624	5.52	8.94	0.49	5.43 g/dl	-0.18	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	634	3.25	9.15	0.30	3.1 g/dl	-0.51	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	665	8.17	14.31	1.17	8.6 mg/dl	0.37	0.09
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	734	108.76	10.65	11.58	98 mg/dl	-0.93	0.85
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	675	145.97	13.24	19.32	139.7 mg/dl	-0.32	1.49

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****

LAB MONTHLY SUMMARY



Lab Name **DAGAR DIAGNOSTIC CENTRE** Lab No **14266**
 Month **December** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 18/12/2021

Date of Report Published : 03/01/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	735	220.59	10.06	22.19	226.2 mg/dl	0.25	1.64
2	UREA II	Urease UV / GLDH II	Transasia / Erba	555	110.52	10.22	11.30	117.5 mg/dl	0.62	0.96
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	488	2.27	9.44	0.21	2.35 mg/dl	0.37	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	649	2.93	11.30	0.33	2.66 mg/dl	-0.82	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	665	5.34	9.22	0.49	5.36 g/dl	0.04	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	650	3.21	9.20	0.30	3.1 g/dl	-0.37	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	637	4.08	14.87	0.61	4.3 mg/dl	0.36	0.05
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	716	104.99	12.01	12.61	97 mg/dl	-0.63	0.94
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	605	85.49	17.53	14.98	74.3 mg/dl	-0.75	1.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JANUARY 2022



Lab Name **DAGAR DIAGNOSTIC CENTRE** Lab No **14266**
Constituent Group **Chemistry II** Date of Result Entered : **19/01/2022**
PT item **Lyophilized Serum** Date of Report Published : **07/02/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	687	130.44	8.67	11.30	141.7 mg/dl	1.00	0.86
2	UREA II	Urease UV / GLDH II	Transasia / Erba	537	47.63	9.86	4.70	51.2 mg/dl	0.76	0.41
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	434	1.70	10.33	0.18	1.67 mg/dl	-0.17	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	615	2.30	15.28	0.35	2.36 mg/dl	0.17	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	556	5.38	9.07	0.49	5.55 g/dl	0.35	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	577	3.15	8.76	0.28	2.98 g/dl	-0.62	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	592	7.17	12.44	0.89	7.5 mg/dl	0.37	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	636	101.83	11.62	11.83	94 mg/dl	-0.66	0.94
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	577	115.89	15.40	17.85	101.9 mg/dl	-0.78	1.49

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
DAGAR DIAGNOSTIC CENTRE
OPP CIVIL HOSPITAL
PALWAL
HARYANA121102

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****