



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 3937

Distribution No.: 152-K

Month/Year: March/2021

Instrument ID: SYSMEX XN-350 SN.11534 02/2016

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 24-05-2021[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	6.02	5.85	11.87	15.4	0.3590	-0.48	0.17	0.13	0.0180	0.25
RBC x10 <sup>6</sup> /µl	1	4.12	4.06	8.18	8.28	0.0130	-0.47	0.06	0.05	0.0040	0.27
Hb g/dl	1	11.9	11.9	23.8	23.9	0.0420	-0.13	0	0.1	0.0110	-1.35
HCT%	1	37.9	36.7	74.6	74.4	0.3300	0.03	1.2	0.4	0.0370	2.40
MCV-fl	1	92	90.4	182.4	179	0.5920	0.33	1.6	0.2	0.0310	3.78
MCH-Pg	1	29.3	28.9	58.2	57.8	0.1200	0.19	0.4	0.2	0.0260	0.90
MCHC-g/dl	1	32.4	31.4	63.8	63.7	0.2690	0.02	1	0.2	0.0280	2.70
Plt. x10 <sup>3</sup> /µl	1	214	205	419	360.5	3.51	1.00	9	9	0.85	0.00
Retic %	2	0.6	0.5	1.1	5	0.15	-1.32	0.1	0.3	0.10	-0.90

### P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbcs=1.00 , Poly=41.00 L=2.00, E=1.00, Mono/Promono=2.00 , B1=2.00 P.M.=4.00, Mye=32.00, Meta=12.00, Other=0.00
RBC Morphology	3	Poly: 30 - 65, Myelo: 10 - 35, Meta: 5 - 20, Promyelo/Blast/Lympho: 1 - 10, nRBC/Baso/Eos/Mono: 0 - 5
Diagnosis	3	CHRONIC MYELOID LEUKEMIA. Chronic Myeloid Leukemia (CML)

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 152--K	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	136	135	85.93	85.93	8.15	8.15	5.92	5.92
<b>RBC x10<sup>6</sup>/µl</b>	1	136	136	88.97	91.18	3.68	3.68	7.35	5.14
<b>Hb g/dl</b>	1	136	136	89.71	89.71	5.88	1.47	4.41	8.82
<b>HCT%</b>	1	136	135	93.33	91.11	5.19	2.96	1.48	5.93
<b>MCV-fl</b>	1	136	135	94.07	87.41	4.44	3.7	1.49	8.89
<b>MCH-Pg</b>	1	136	135	91.85	91.11	3.7	2.96	4.45	5.93
<b>MCHC-g/dl</b>	1	136	135	94.07	86.67	4.44	5.19	1.49	8.14
<b>Plt. x10<sup>3</sup>/µl</b>	1	136	134	90.3	87.31	5.97	5.22	3.73	7.47
<b>ReticCount%</b>	2	136	136	83.09	75.74	2.94	0.74	13.97	23.52
<b>PS Assessment</b>	3	136	123	Satisfactory :92, Borderline Sat. :4, Unsatisfactory :4					

**\*Comments:**

1). **Among Lab (EQA) : Results acceptable.**

2). **Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 3937

Distribution No.: 153-K

Month/Year: October/2021

Instrument ID: SYSMAX XN-350 (11534 02/2016)

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 10-12-2021[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	4.46	4.37	8.83	8.7	0.0440	0.13	0.09	0.1	0.0100	-0.08
RBC x10 <sup>6</sup> /µl	1	5.25	5.17	10.42	10.4	0.0130	0.06	0.08	0.05	0.0040	0.51
Hb g/dl	1	16.3	16	32.3	32.5	0.0460	-0.19	0.3	0.1	0.0100	1.35
HCT%	1	54.1	53.9	108	99.6	0.2350	1.30	0.2	0.5	0.0390	-0.58
MCV-fl	1	104.3	103	207.3	190.7	0.3630	1.73	1.3	0.35	0.0290	1.83
MCH-Pg	1	31.5	31	62.5	62.4	0.0880	0.05	0.5	0.25	0.0190	0.84
MCHC-g/dl	1	30.2	30.1	60.3	65.35	0.1640	-1.20	0.1	0.3	0.0210	-0.67
Plt. x10 <sup>3</sup> /µl	1	150	141	291	284	1.84	0.14	9	5	0.36	0.70
Retic %	2	3.5	3	6.5	7	0.14	-0.14	0.5	0.3	0.02	0.67

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=0.00 , Poly=4.00 L=94.00, E=0.00, Mono/Promono=1.00 , B1=1.00 P.M.=0.00, Mye=0.00, Meta=0.00, Other=0.00	Lymp: 85-94, Poly: 4-12, blast: 1-8, nRBC/mono/Eosino/Myelo/Meta: 0-1		
RBC Morphology	3	normocytic normochromic with few microcytes and spherocytes	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, Microcytosis, Hypochromia; Mild: Macrocytosis, Poikilocytosis		
Diagnosis	3	CLL	Chronic Lymphocytic Leukemia (CLL)		

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 153--K	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	267	267	77.9	88.01	6.74	2.25	15.36	9.74
<b>RBC x10<sup>6</sup>/µl</b>	1	267	267	91.01	86.52	6.37	7.49	2.62	5.99
<b>Hb g/dl</b>	1	267	267	88.39	61.42	7.87	32.21	3.74	6.37
<b>HCT%</b>	1	267	266	95.86	87.97	2.26	5.26	1.88	6.77
<b>MCV-fl</b>	1	267	266	94.36	95.49	4.89	2.26	0.75	2.25
<b>MCH-Pg</b>	1	267	266	88.35	89.47	7.14	3.38	4.51	7.15
<b>MCHC-g/dl</b>	1	267	266	95.49	89.85	3.38	4.14	1.13	6.01
<b>Plt. x10<sup>3</sup>/µl</b>	1	267	266	95.11	90.98	3.38	4.89	1.51	4.13
<b>ReticCount%</b>	2	267	267	95.13	84.27	2.62	1.5	2.25	14.23
<b>PS Assessment</b>	3	267	232	Satisfactory :93.14%, Borderline Sat. :0.86%, Unsatisfactory :6%					

**\*Comments:**

1). **Among Lab (EQA) : Results acceptable.**

2). **Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between " $0$  to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value ( $0.3*SDPA$ ). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value ( $0.3*SDPA$ ).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----