



Innolyte
Fastest to Analyze

Calibration Certificate

Hospital Name	:	Nimaya Diagnostics Trivandrum
Address	:	Main Road, Kunnathukal, Karakonam P O Trivandrum
Instrument Name	:	Innolyte
Serial No.	:	0350911161486
Department	:	Laboratory
Calibration Date	:	17-11-2021
Next Calibration Due	:	17-11-2022
Appearance	:	Clean

Calibration Status:

Item	Reagent	Value (mV)	Range	Remarks
Empty Tube ADC	Air	323	0-500	OK
Calibration A ADC	STD A	1235	>800	OK
Valve to Sensor Steps	STD A	1286	<2000	OK
Sodium	STD A	77.21	Stable	Ok
Sodium	STD B	71.45	Stable	Ok
Potassium	STD A	70.43	Stable	Ok
Potassium	STD B	85.6	Stable	Ok

Item	Concentration		
	Sodium	Potassium	Remarks
STD A	140	4	OK
STD B	110	8	OK

This is to certify that above instrument is calibrated and validated as per specification by the manufacturer and instrument is ready to report the samples of the patients with daily quality controls validation.

Performed by,

Amal V K
Service Engineer
Ernakulam