

Registered Office:
BioSystems Diagnostics Pvt. Ltd.,
A4, SIPCOT Industrial Park, Irungattukottai,
Thandalam Village, Sriperumbudur - 602105.
Ph: +91 - 44 - 3034 1911 / 12 / 13
E-mail: customersupport@biosystems.in
Website: www.biosystems.in
CIN - U29297TN2003PTC051492



Date: 21/06/2021

CALIBRATION CERTIFICATE

This is to certify that BTS – 350 Photometer bearing **S.No: 801751324** installed at **Prem Clinical Lab,Cuddalore** has been calibrated on 21/06/2021.

Calibration Values: LIGHT INTENSITY TEST

LED	TINT	MAIN	REFERENCE
340	95	964159	290503
405	140	976943	843661
505	69	911963	970894
535	67	967907	825315
560	98	910798	970322
600	39	963602	770829
635	30	953995	706093
670	119	787900	977612

PERISTALTIC PUMP ADJUSTMENT TEST

VALUE	CURRENT	NEW
SAMPLE VOLUME	53449	53449
POSITIONING	2200	2200
PUMP DELAY	2	2

Next calibration is due on 20/06/2022

For **BIOSYSTEMS DIAGNOSTICS Pvt. Ltd.**,

A. Prayadharan

AUTHORISED SIGNATORY

