



CERTIFICATE OF CALIBRATION/ VERIFICATION

Customer Name: AVIOR CLINICAL DIAGNOSTIC CENTRE
 Model: Fully Automated Bio-Chemistry Analyzer-EM 200
 Serial No: B160215
 Calibration Date: 27-12-2021
 Next Calibration due on before 26-12-2022
 Calibrator Used: Multical Lot Number: 2105079
 Quality Control Used: ERBA NORM CONTROL S071915A

This is to certify that the above mentioned instrument has been verified of calibration according to the procedures provided by the manufacturer.

General Parameters:

Parameter	Measured Value	Acceptable Range	
Line Voltage	230V	230 V +/- 5%	Pass
Logic Voltage	5.0 V	4.9 V to 5.1 V	Pass
Lamp Voltage	12.0 V	11.8 V to 12.2 V	Pass
Flow cell Temperature	37.0 deg	37 +/- 0.5	Pass
45 Cuvette blank value for for all filter	Enclose valuesheet for all cuvette	Abs 0.02-0.2	Pass

Photometer Filter Gain Levels:

Wavelength	Offset	Gain
340	255	560
405	255	507
505	255	518
546	255	506
578	255	485
600	255	478
660	255	468
700	255	460

Avit
27/12/21

Technical Service Department

Trasasia Bio-Medical Ltd,

RSM.