

OFFICE OF THE MEDICAL OFFICER I/C ICTC-2 DIST. HEAD. QUATER HOSPITAL ,
DHENKANAL

This is to certified that one domestics refrigerator of Samsung make (stock entry in page no- 171) in fixed Asset stock register of ICTC-2 is used in ICTC LAB Dist. Hqrs. Hospital , Dhenkanal is functioning well and It is maintaining the temperature properly (i.e 2-8^oC) as per daily temperature log sheet .

Bachant Nath Saha
Lab Technician

ICTC-2 DHH Dhenkanal

[Signature]
Medical officer I/C

ICTC-2 DHH Dhenkanal

Medical Officer I/C
ICTC-2, Dist. Hqrs. Hospital
Dhenkanal

Annexure A7: Temperature log book at HCTS Facilities

National AIDS Control Organization Daily Temperature Log Sheet

(HIV Counselling & Testing Services Facility)

Name of HCTS Facility (SA-ICTC/F-ICTC/PPP/TTI etc): ICTC-2
 Address: Dist. Hqrs. Hospital, Dhenkanal
 Name of the Staff: Debmalya Pradhan
 Designation: Lab. Technician Contact No: 9861003281
 Name of In-charge of HCTS Facility: Dr. K. Nayak M.S. (P&G)
 Designation: Medical officer IIC Contact No: _____

Month: January Year: 2022

Day	AM		PM		Remarks
	Time	Temperature (in °C)	Time	Temperature (in °C)	
1	9.00	4.8	4.00	4.0	
2	9.15	4.8	5.00	5.0	
3	9.20	5.0	5.00	5.0	
4	9.00	4.8	5.10	5.0	
5	9.00	3.8	4.00	5.2	
6	9.00	5.0	4.30	5.2	
7	9.00	4.0	4.30	5.0	
8	9.30	4.6	4.50	4.8	
9	9.05	4.6	4.45	5.0	
10	9.10	4.8	4.00	5.0	
11	9.10	5.0	4.00	5.0	
12	9.10	6.0	4.15	5.0	
13	9.00	6.2	4.15	5.0	
14	9.30	5.0	4.00	5.0	
15	9.20	5.0	4.00	5.0	
16	9.00	5.0	4.00	5.0	
17	9.00	5.0	4.00	5.0	
18	9.05	5.2	5.00	5.2	
19	9.10	5.3	5.15	5.4	
20	9.15	5.4	5.00	5.4	
21	9.20	5.0	5.00	5.0	
22	9.20	5.0	5.15	5.0	
23	9.00	5.4	5.20	5.0	
24	9.00	5.0	5.00	5.0	
25	9.00	6.0	4.30	5.0	
26	9.30	5.4	4.30	5.0	
27	9.30	5.4	4.45	5.0	
28	9.00	5.8	4.20	5.0	
29	9.10	5.8	4.20	4.8	
30	9.10	5.8	4.00	4.8	
31	9.00	6.0	5.00	5.0	

Debmalya Pradhan
 Lab-Tech
 ICTC-2, Dist. Hqrs. Hospital
 Dhenkanal

Dr. K. Nayak
 Medical Officer IIC
 ICTC-2, Dist. Hqrs. Hospital
 Dhenkanal

Annexure A7: Temperature log book at HCTS Facilities

**National AIDS Control Organization
Daily Temperature Log Sheet
(HIV Counselling & Testing Services Facility)**

Name of HCTS Facility (SA-ICTC/F-ICTC/PPP/TL etc): ICTC-2
 Address: Dist. Hqrs. Hospital, Dhenkanal
 Name of the Staff: Pradyumn Nath Sahu
 Designation: Lab. Technician Contact No: 986103281
 Name of In-charge of HCTS Facility: Dr. R. Nayak M.S. (O&A)
 Designation: Medical Officer II Contact No: _____

Month: December Year: 2021

Day	AM		PM		Remarks
	Time	Temperature (in °C)	Time	Temperature (in °C)	
1	9.00	5.0	5.00	6.0	
2	9.15	4.8	4.45	6.0	
3	9.20	5.0	4.30	5.0	
4	9.30	4.8	4.00	4.8	
5	9.30	4.8	4.00	4.0	
6	9.15	4.6	4.05	6.0	
7	9.15	5.0	4.10	6.0	
8	9.20	5.0	4.05	5.0	
9	9.20	5.0	4.00	5.0	
10	9.30	5.0	4.10	6.0	
11	9.30	4.8	4.15	5.0	
12	9.00	4.6	4.20	4.0	
13	9.00	4.6	4.20	4.8	
14	9.00	4.3	4.30	5.0	
15	9.00	4.0	4.30	5.0	
16	9.00	4.6	4.15	5.0	
17	9.00	5.0	4.20	5.0	
18	9.00	5.0	4.20	5.0	
19	9.15	5.0	4.20	5.0	
20	9.30	5.0	4.30	5.0	
21	10.00	5.0	4.30	5.0	
22	9.45	5.0	4.45	5.0	
23	9.00	5.0	4.45	4.0	
24	9.00	5.0	5.00	6.0	
25	9.05	5.0	5.00	6.0	
26	9.00	5.0	5.00	6.0	
27	9.00	5.0	4.45	5.0	
28	9.20	4.8	5.15	5.0	
29	9.20	6.0	5.15	5.0	
30	9.30	6.2	5.15	4.0	
31	9.05	6.4	5.30	6.0	

[Handwritten Signature]

[Handwritten Signature]
 Lab. Tech
 ICTC-2, Dist. Hqrs. Hospital
 Dhenkanal

Medical Officer IIC
 ICTC-2, Dist. Hqrs. Hospital
 Dhenkanal