

Date: 21/06/2021

CALIBRATION CERTIFICATE

This is to certify that BTS – 350 Photometer bearing S.No: **801751324** installed at **Prem Clinical Lab,Cuddalore** has been calibrated on 21/06/2021.

Calibration Values: LIGHT INTENSITY TEST

LED	TINT	MAIN	REFERENCE
340	95	964159	290503
405	140	976943	843661
505	69	911963	970894
535	67	967907	825315
560	98	910798	970322
600	39	963602	770829
635	30	953995	706093
670	119	787900	977612

PERISTALTIC PUMP ADJUSTMENT TEST

VALUE	CURRENT	NEW
SAMPLE VOLUME	53449	53449
POSITIONING	2200	2200
PUMP DELAY	2	2

Next calibration is due on 20/06/2022

For **BIOSYSTEMS DIAGNOSTICS Pvt. Ltd.,**

A. P. S. Yadav


AUTHORISED SIGNATORY

Registered Office:

BioSystems Diagnostics Pvt. Ltd.,
A4, SIPCOT Industrial Park, Irungattukottai,
Thandalam Village, Sriperumbudur - 602105.
Ph: +91 - 44 - 3034 1911 / 12 /13
E-mail: customersupport@biosystems.in
Website: www.biosystems.in
CIN - U29297TN2003PTC051492



BTS 350 - Semi Auto BioChemistry Analyser

Installation Qualification Certificate

For

***PREM CLINICAL LAB,
CUDDALORE***

BioSystems Diagnostics Pvt. Ltd

Corporate Office:
BioSystems S.A. Costa Brava 30, 08030 Barcelona (Spain)
Tel +34-93 311 00 00, Fax+34-93 346 77 99
<http://www.biosystems.es>



Registered Office:
BioSystems Diagnostics Pvt. Ltd.,
A4, SIPCOT Industrial Park, Irungattukottai,
Thandalam Village, Sriperumbudur - 602105.
Ph: +91 - 44 - 3034 1911 / 12 /13
E-mail: customersupport@biosystems.in
Website: www.biosystems.in
CIN - U29297TN2003PTC051492



CERTIFICATE OF INSTALLATION

Instrument Name : BTS 350 Semi Auto Biochemistry Analyser

Serial Number : 801751324

**Customer : Prem Clinical Lab,
No. 82-c, Main Road,
Pudupalayam, Cuddalore.607001.**

The undersigned performer certifies that the installation Qualification Protocol has been successfully completed for the instrument stated above.

Engineer

Signature :
Name : Mr.Ramesh.G
Designation : Asst Manager - Service
Company : BioSystems Diagnostics Pvt. Ltd

Quality Manager

Signature :
Name : R.JANAKIRAMAN
Designation : Quality Manager
Company : Prem Clinical Lab, Cuddalore.



Registered Office:

BioSystems Diagnostics Pvt. Ltd.,
A4, SIPCOT Industrial Park, Irungattukottai,
Thandalam Village, Sriperumbudur - 602105.
Ph: +91 - 44 - 3034 1911 / 12 /13
E-mail: customersupport@biosystems.in
Website: www.biosystems.in
CIN - U29297TN2003PTC051492



INSTALLATION REPORT

Customer Name : PREM CLINICAL LAB, CUDDALORE
Department : Quality Manager
Contact Person : R.JANAKIRAMAN
Instrument Model : BTS 350
Serial Number : 801751324
Date of Installation : 10/12/2010

The System was installed along with the necessary accessories. All basic requirements and adjustments were checked and are found to be satisfactory. Preliminary customer training was provided and standardization of a few parameters was done. The Instrument is found to be working satisfactorily.



Registered Office:

BioSystems Diagnostics Pvt. Ltd.,
A4, SIPCOT Industrial Park, Irungattukottai,
Thandalam Village, Sriperumbudur - 602105.
Ph: +91 - 44 - 3034 1911 / 12 /13
E-mail: customersupport@biosystems.in
Website: www.biosystems.in
CIN – U29297TN2003PTC051492



INSTRUMENT SETUP:

- Instrument installed
- Printer Installed
- User software 1.42 installed
- All necessary checks done

OPERATION INSPECTION:

- No air bubbles found in the Tubing
- Temperature of the Flow cell Chamber are within range
- No unusual noises produced in the system
- Warm up procedure within time limit

For Biosystems Diagnostics Pvt. Ltd.,

For

A. Pulyadhi



Mr. Ramesh.G
Asst Manager – Service



Registered Office:
BioSystems Diagnostics Pvt. Ltd.,
A4, SIPCOT Industrial Park, Irungattukottai,
Thandalam Village, Sriperumbudur - 602105.
Ph: +91 - 44 - 3034 1911 / 12 / 13
E-mail: customersupport@biosystems.in
Website: www.biosystems.in
CIN - U29297TN2003PTC051492



INSTALLATION QUALIFICATION

Customer Name : PREM CLINICAL LAB, CUDDALORE

Address : PREM CLINICAL LAB,
No. 82-c, Main Road,
Pudupalayam, Cuddalore.607001.

Instrument Model : BTS 350

Serial No. : 801751324

Operational requirement for the Instrument:

Requirement	Acceptable Range	Observed Range
Input Voltage	220 \pm 10 V	222 V
Line Frequency	50 or 60 Hz	50 Hz
Ambient Temperature	18 – 30 deg. C	23 deg. C

For Biosystems Diagnostics Pvt. Ltd.,

Pon
A. P. Pradyumnachandran

Mr. Ramesh. G
Asst Manager - Service



Registered Office:

BioSystems Diagnostics Pvt. Ltd.,
A4, SIPCOT Industrial Park, Irungattukottai,
Thandalam Village, Sriperumbudur - 602105.
Ph: +91 - 44 - 3034 1911 / 12 /13
E-mail: customersupport@biosystems.in
Website: www.biosystems.in
CIN – U29297TN2003PTC051492



BTS350 - Semi Auto BioChemistryAnalyser

Operational Qualification

For

***PREM CLINICAL LAB,
CUDDALORE***

BioSystems Diagnostics Pvt. Ltd.,

Corporate Office:
BioSystems S.A. Costa Brava 30, 08030 Barcelona (Spain)
Tel +34-93 311 00 00, Fax+34-93 346 77 99
<http://www.biosystems.es>



Registered Office:

BioSystems Diagnostics Pvt. Ltd.,
A4, SIPCOT Industrial Park, Irungattukottai,
Thandalam Village, Sriperumbudur - 602105.
Ph: +91 - 44 - 3034 1911 / 12 /13
E-mail: customersupport@biosystems.in
Website: www.biosystems.in
CIN – U29297TN2003PTC051492



BTS350 – Semi Auto Biochemistry Analyser

OPERATIONAL QUALIFICATION

1. System Decontamination
2. Maintenance Procedures
3. System Performance check procedures
4. Customer Training Operation and Maintenance



Registered Office:
BioSystems Diagnostics Pvt. Ltd.,
A4, SIPCOT Industrial Park, Irungattukottai,
Thandalam Village, Sriperumbudur - 602105.
Ph: +91 - 44 - 3034 1911 / 12 /13
E-mail: customersupport@biosystems.in
Website: www.biosystems.in
CIN - U29297TN2003PTC051492



CERTIFICATE OF OPERATION

Instrument Name : BTS350 - Semi Auto Biochemistry Analyser
Serial Number : 801751324
Customer : PREM CLINICAL LAB,
No. 82-c, Main Road,
Pudupalayam, Cuddalore.607001.

The undersigned performer certifies that the Operational Qualification Protocol has been successfully completed for the instrument stated above.

Engineer

For

A. P. Ramesh G



Signature :
Name : Mr. Ramesh. G
Designation : Asst Manager - Service
Company : BioSystems Diagnostics Pvt. Ltd.,

Quality Incharge

Signature :
Name : R. JANAKIRAMAN
Designation : Quality Manager
Company : Prem Clinical Lab, Cuddalore



Registered Office:

BioSystems Diagnostics Pvt. Ltd.,
A4, SIPCOT Industrial Park, Irungattukottai,
Thandalam Village, Sriperumbudur - 602105.
Ph: +91 - 44 - 3034 1911 / 12 /13
E-mail: customersupport@biosystems.in
Website: www.biosystems.in
CIN – U29297TN2003PTC051492



SYSTEM PERFORMANCE CHECK PROCEDURE

SYSTEM CHECK	PERFORMANCE
System Decontamination	Done
Photometer LED Check	Done
Start Up	Done
Pump Adjustments	Done
Temperature Check	Done

Customer Signature:

Corporate Office:
BioSystems S.A. Costa Brava 30, 08030 Barcelona (Spain)
Tel +34-93 311 00 00, Fax+34-93 346 77 99
<http://www.biosystems.es>



Registered Office:
BioSystems Diagnostics Pvt. Ltd.,
A4, SIPCOT Industrial Park, Irungattukottai,
Thandalam Village, Sriperumbudur - 602105.
Ph: +91 - 44 - 3034 1911 / 12 /13
E-mail: customersupport@biosystems.in
Website: www.biosystems.in
CIN - U29297TN2003PTC051492



OPERATIONAL QUALIFICATION

System Certification:

Study data has determined that the System described in this document either meets all criteria outlined in this Operational Qualification Protocol, or exceptional conditions have been identified and documentation included. Exceptional conditions, if any have been addressed. The system is ready for specified usage.

Protocol Performed By: BioSystems Diagnostics Pvt. Ltd
Name : Mr.Ramesh.G
Title : Asst Manager - Service
Company: BioSystems Diagnostics Pvt. Ltd

Customer Authorization PREM CLINICAL LAB, CUDDALORE
Name:R.JANAKIRAMAN
Title: Quality Manager

Pon
A. Pulyadhisaran

ENGINEER SIGNATURE

CUSTOMER SIGNATURE

DATE:

DATE:

Corporate Office:
BioSystems S.A. Costa Brava 30, 08030 Barcelona (Spain)
Tel +34-93 311 00 00, Fax+34-93 346 77 99
<http://www.biosystems.es>



Registered Office:

BioSystems Diagnostics Pvt. Ltd.,
A4, SIPCOT Industrial Park, Irungattukottai,
Thandalam Village, Sriperumbudur - 602105.
Ph: +91 - 44 - 3034 1911 / 12 /13
E-mail: customersupport@biosystems.in
Website: www.biosystems.in
CIN - U29297TN2003PTC051492



BTS350 - Semi Auto BioChemistryAnalyser

Performance Qualification

For

***PREM CLINICAL LAB,
CUDDALORE***

BioSystems Diagnostics Pvt. Ltd

Corporate Office:
BioSystems S.A. Costa Brava 30, 08030 Barcelona (Spain)
Tel +34-93 311 00 00, Fax+34-93 346 77 99
<http://www.biosystems.es>



BTS350 SEMIAUTO BIOCHEMISTRY ANALYSER

PERFORMANCE QUALIFICATION:

The following parameters were checked and found to be within acceptable limits.

<u>LED/Filters</u>	<u>Sensitivity (nA)</u> (Observed values)	<u>Minimum Value (nA)</u>	<u>Maximum Value (nA)</u>
340	30	12	40
405	25	12	60
505	23	12	100
535	60	12	100
560	173	150	450
600	23	12	100
635	12	12	30
670	32	12	120

DAC Analogic/Digital Converter

<u>Filters</u>	<u>Sensitivity (nA)</u> (Observed values)	<u>Minimum Value (nA)</u>	<u>Maximum Value (nA)</u>
340	67	50	90
405	95	75	140
505	98	75	140
535	88	75	140
560	117	60	120
600	78	75	140
635	92	75	140
670	84	75	140

