



Date: 09.02.2022

## Certificate of Calibration

<u>Customer Name:</u> Christian Hospital Sewa Sansthan (CH PHARMACY LAB) Harbanshpur, Azamgarh

Model: Semi-Automated Bio Chemistry Analyzer Erba Chem 5 Plus V2.

Serial No.: S190616

Calibration done date: 09.02.2022

Next Calibration due Date: 08.02.2023

Lab In charge: Miss. Srishti Ram

This is to certify that above mentioned product has been verified of calibration By

Checking Lamp Voltage

Checking gain of Photo meter for all Filters

Checking sipping volume of peristaltic pump

Filter	Offset	Gain
340	-3528	118884
405	-3335	131262
450	-3395	124185
505	-3326	116761
546	-3237	121074
578	-3182	118150
600	-3368	119188
670	-3221	121853

Followed by calibration of all the three types of chemistries (picked up CHOL, CREAT and SGPT) and QC run with satisfactory values of QC

Calibration at site performed by Mr. Sarvesh Singh

Authorized Signatory

(Mr. Sarvesh Singh) Sr. Service Engineer

For Transasia Bio-Medicals Ltd.