



YEA

☎ : 033 2677 1412, 2667 1909, 82329 85881
Tele Fax : 033 2677 1411

YOUNG ENGG. & CALIBRATION SERVICES PVT. LTD.

Admn. Office & Lab : 44, Shahid Khudiram Bose Sarani, (formerly Kamardanga Road), Ichapur, Howrah - 711 104
E-mail : calibration@yeagroup.org ★ Visit us at : www.yeagroup.org
CIN : U74140WB2005PTC106389 ★ Toll Free Help Desk : 1800 345 9001

Calibration Certificate

Certificate No.: YEA(C)/ 220205SA/S2970/003

Dated : 07.02.2022

Calibrated for : State Referral Hospital & Zoram Medical Hospital, ICTC
Falkawn, Mizoram,
Pin - 796005.

Calibrated on.	Next Calibration Due On.	Page No.	No. of Pages
05.02.2022	05.02.2023	1	2

Description of Instrument

<i>Name</i>	: Centrifuge
<i>Identification</i>	: SRHF/ICTC/CNF-61
<i>Make</i>	: Remi (Sl. No. :BFLC-1618)
<i>Model / Type</i>	: C852

Service Request No. / Order Reference : 220205SA/S2970/001-003

Receipt of Instrument : 05.02.2022

Location where calibration performed : At site

Applicable Specification of Instrument to be Calibrated : Not Specified

Condition of Instrument to be Calibrated : Satisfactory

Basis of Calibration : Comparison Method with Standard

Environmental Condition during Calibration : Temp. : 23±2°C RH 50±10%

Equipment used for Calibration : Standard Digital Tachometer

Traceability : To National Standard vide Certificate No. HF/CC/210909-04/001 dated 14.09.2021

Measurement Uncertainty : The figures are stated in the result at coverage factor $k = 2$ corresponding to confidence probability of 95% for normal distribution.

Laboratory Identification : Sticker is affixed for authentication of calibration

The above Instrument/Equipment/Machine has been calibrated up to its range.
The results are given in the page 2 to ... 2 ... of the certificate

YEA(C)F-08





YEA

☎ : 033 2677 1412, 2667 1909, 82329 85881

Tele Fax : 033 2677 1411

YOUNG ENGG. & CALIBRATION SERVICES PVT. LTD.

Admn. Office & Lab : 44, Shahid Khudiram Bose Sarani, (formerly Kamardanga Road), Ichapur, Howrah - 711 104

E-mail : calibration@yeagroup.org ★ Visit us at : www.yeagroup.org

CIN : U74140WB2005PTC106389 ★ Toll Free Help Desk : 1800 345 9001

Certificate No.: YEA(C)/220205SA/S2970/003

Calibrated By : S. Ghosh

Dated : 07.02.2022

Page 2 of 2 Pages

Calibration Results

CALIBRATION RESULTS OF CENTRIFUGE :

Knob Set At	Standard Value	Overall Measurement Uncertainty
	rpm	± rpm
1	781.3	3.8
2	1753.9	
3	2251.5	
4	2868.2	
5	3476.5	

DUC : Devices Under Calibration

REMARKS : The above Centrifuge has been calibrated and the readings are tabulated above.

NOTES : This certificate refers to the results tabulated at the time of calibration only for the above mentioned item submitted for calibration and this should not be reproduced, other than in full except with prior written approval from the head of calibration laboratory.

Authorised By

A. Das

Manager Calibration

*** End of the Calibration Certificate ***

