

VIEW LAB MONTHLY SUMMARY

Lab Name: WELLNESS DIAGNOSTIC PVT LTD
 Lab No: 10041
 Month: December
 Year: 2021
 Constituent Group: Chemistry I

Click on the analyte to view Graphical Data

Date of Result Entered : 2012/2021

Date of Report Published : 03/01/2022

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE	GOD-POD	Mindray	196	191.38	4.85	9.28	194.1 mg/dl	0.29	1.33
2	UREA	Urease UV / GLDH	Mindray	191	87.91	6.88	6.05	96.8 mg/dl	1.47	0.88
3	CREATININE	Jaffes End point	Any Analyser (Automation / Semi Automation)	387	6.55	12.58	0.82	7.3 mg/dl	0.91	0.08
4	TBILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Mindray	165	2.36	10.94	0.26	2.46 mg/dl	0.39	0.04
5	T-PROTEIN	Biuret - Colorimetric	Mindray	197	5.21	6.28	0.33	5.1 g/dl	-0.34	0.05
6	ALBUMIN	BCG - colorimetric	Mindray	193	3.29	8.22	0.27	3.1 g/dl	-0.70	0.04
7	CALCIUM	Arsenazo III	Mindray	141	10.60	6.60	0.70	10.5 mg/dl	-0.14	0.12
8	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Mindray	103	4.87	10.28	0.48	4.25 mg/dl	-0.88	0.09
9	URIC ACID	Enzymatic / Uricase Colorimetric	Mindray	194	8.69	8.08	0.70	8.53 mg/dl	-0.23	0.10
10	CHOLESTEROL	CHOD-PAP	Mindray	203	108.30	5.93	6.42	111.4 mg/dl	0.48	0.90
11	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Mindray	181	209.18	5.82	12.17	211.9 mg/dl	0.22	1.81
12	HDL CHO	Direct method / Enzymatic colorimetric	Mindray	144	24.96	10.71	2.67	24.9 mg/dl	-0.02	0.45
13	AST	UV Kinetic without PLP (P-5-P)	Mindray	140	92.12	9.47	8.73	88 U/L	-0.47	1.48
14	ALT	UV Kinetic without PLP (P-5-P)	Mindray	132	59.18	14.79	8.75	47.2 U/L	-1.37	1.52
15	ALP	PNP AMP Kinetic	Mindray	123	236.91	14.88	35.26	386.3 U/L	4.24	6.36
16	IRON	Ferro Zine (No Protein Removal)	Any Analyser (Automation / Semi Automation)	46	89.69	12.90	11.57	78 ug/dl	-1.01	3.41
17	MAGNESIUM	Others (Any other principles / Methods)	Any Analyser (Automation / Semi Automation)	120	2.49	15.45	0.38	2.91 mg/dl	1.09	0.07

SDI Range Interpretation

Lab Name WELLNESS DIAGNOSTIC PVT LTD
 Lab No 10041
 Month December
 Year 2021
 Constituent Group HbA1c

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

Date of Result Entered : 20/12/2021

Date of Report Published : 03/01/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	HPLC-Ion Exchange	Tosoh G8 - HPLC / HPLC-723G8)	99	5.68	5.63	0.32	5.5 %	-0.56	0.06

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out