

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

OLIVE HEART AND DIAGNOSTICS (601)

Outlier And Analyte Summary Report

Result Details For Cycle No 210512 and Sample No 01


Report Date : 30/12/2021

Instrument : Med Source Csense 100

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	205.66	196.73	-0.24	-0.23
@ Urea	mg/dl	73.10	73.10	-0.33	-0.11
@ Creatinine	mg/dl	3.43	3.87	-1.48	-0.93
@ Bilirubin Total	mg/dl	3.73	3.57	0.99	0.85
@ Bilirubin Direct	mg/dl	0.99	0.96	-0.30	0.14
@ Total Protein	g/dL	3.41	3.41	0.18	-0.70
@ Albumin	g/dL	2.13	2.27	-0.54	-0.64
@ Uric Acid	mg/dl	6.34	6.27	-0.97	-0.72
@ Cholesterol	mg/dl	83.23	82.12	0.67	0.24
@ Triglycerides	mg/dl	58.01	60.26	-1.91	-0.18
@ HDL	mg/dl	26.19	21.10	1.30	-0.46
@ ALT	U/L	66.97	63.49	0.09	0.26
@ AST	U/L	145.72	150.61	-0.06	0.52
@ ALP	U/L	324.07	392.04	-1.31	-0.80

Legend @ : z score ≤ 2.0 - Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score ≥ 3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	14
Not Evaluated Parameters	0
Evaluated Parameters	14
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A,
 Kohinoor Mall, Kiroi Road, Kurla (W),
 Mumbai - 400070

End of Report