

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**OLIVE HEART AND DIAGNOSTICS (601)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 210512 and Sample No 01**


**Report Date : 30/12/2021**

**Instrument : Sysmex XP-100**

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	5.13	5.07	0.14	0.29
@ Total Leucocytes (WBC) count	/c.mm	4900.00	4,611.86	0.56	1.71
@ Haemoglobin	g/dL	16.20	16.20	-0.44	-0.26
@ MCV (Mean Corpuscular Volume)	fL	103.40	99.89	0.62	0.39
@ MCH (Mean Corpuscular Hb)	pg	31.55	31.95	-0.34	-0.33
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	30.50	31.66	-0.71	-0.69
@ Platelet Count	1000/ $\mu$ L	144.00	163.95	-1.06	-1.10
@ Hematocrit	%	53.15	50.76	0.92	0.76
@ RDW CV%	%	14.50	14.48	0.03	-0.10

Legend @ : z score  $\leq$  2.0 - Acceptable  
! : 2.0 < z score < 3.0 - Warning  
X : z score  $\geq$  3 - Unacceptable  
# : Not Evaluated  
🕒 : Delayed Result Entry  
\* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



**Dr Puneet Kumar Nigam**

**PT coordinator & Technical Manager, MHL EQAS**

Unit No. 409-416.

Commercial Building - 1A,

Kohinoor Mall, Kiroi Road, Kurla (W),

Mumbai - 400070

End of Report