



CERTIFICATE OF CALIBRATION

COMMITTED TO THE CUSTOMER SINCE - 1996

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.

Contact : 080-23377266, Mob : 9986586789 / 9448080177 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

1 Name and Address of the Customer : M/s. Pooja Diagnostics,
4th Main Rd, Basaveshwara HBCS Layout, 2nd Stage,
Chandra Layout, Attiguppe, Bangalore - 560 040.

2 Customer Reference

2.1 ULR No : CC247322100007648F
2.2 SRF No : 7449
2.3 Certificate No. : VI/21-22/7449-01
2.4 Format No. : VI-FRM-ME-105
2.5 DC No & Date : 001 & 23 - 02 - 2022
2.6 Date Of Issue : 25 - 02 - 2022

3 Details Of Device Under Calibration(DUC).

3.1 Nomenclature : Micropipette
3.2 Make : VERTEX
3.3 Range : 5 - 50 µl
3.4 SL No. : 060210013
3.5 DUC Condition : Satisfactory
3.6 Calibration Procedure No. : SOP-16-71 Based On ISO 8655-6:2002 E
3.7 No.of Pages : 2
3.8 Calibration Date : 24-02-2022
3.9 Calibration Due : 23-02-2023
3.10 Calibration done at : VI Volumetric Lab
3.11 Discipline : Mechanical (Mass & volume)

4 Environmental Condition

Temperature 20.1 °C Humidity 53 %Rh

5 Standards Used for calibration

Sl. No.	Nomenclature	Make & Model	Sl. No	Range	Traceable to /Cert. No.	Validity
1	Digital Weighing Balance	RADWAG/MYA5.4Y	544953	0.1mg to 5g	VI/21-22/INT-ME-121	17-09-2022
2	E1 Class Weights	LCGC	VI-E1-ME-001	1mg to 200g	TVCSPL 19/04/339-01	02 - 04 - 2022

6 Note:

- 6.1. The Calibration Certificate relates only to the above DUC
- 6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab..
- 6.3. Corrections/erasing, invalidate the Calibration Certificate.
- 6.4. Calibration of the DUC are traceable to National standards/International Standards
- 6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6. Results Reported are valid at the time of and under the stated conditions of measurements.
- 6.7. The usage of NABL symbol is as per NABL guidelines given in NABL-133.

Calibrated By

Ranjith Kumar P
Ranjith Kumar P

(Calibration Engineer)

Checked By

P. Santhosh Kumar
P. Santhosh Kumar

(Lab- In-Charge)





COMMITTED TO THE
CUSTOMER SINCE - 1996

CERTIFICATE OF CALIBRATION

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.

Contact : 080-23377266, Mob : 9986586789 / 9448080177 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No: VI/21-22/7449-01

Page 2 of 2

Range: 5 - 50 µl

Results:

Range	Cal Point	Volume observed	Expanded Uncertainty ±
5 - 50 µl	µl	µl	µl
	5.0	4.734	0.02
	20.0	19.798	0.04
	50.0	49.852	0.20

Note:

1. Visual Inspection : Found Well.
2. Tripple Distilled water is used to Calibrate the MicroPipette.


Conclusion /Remarks:

1. Ref. standard used are traceable to National/International Standard
2. The Expanded Uncertainty of associated with measurement at approximate **95.45%** confidence level with coverage factor k=2

Calibrated By


Ranjith Kumar P
(Calibration Engineer)

Checked By


P. Santhosh Kumar
(Lab- In-Charge)

