



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

Customer name And Address	M/S: Wellness Diagnostic Pvt. Ltd. Dharamveer Market, 2 nd Floor Near Seble Cinema Mathura Road Badarpur, New Delhi-110044.	ULR No.	CC223922000000765 F
		Service request No & date	P/02/21/02/2022
		Date of calibration	21/02/2022
		Valid up to	20/02/2023
		Date of issue	22/02/2022

Instrument Details

Instrument name	Refrigerator	S. No.	-----
Make	LG.	Location	Lab
Range / Size	2 to 8 °C	Accuracy	-----
Least Count	1 °C	Visual Inspection	OK
I.D. No.	RF-01	Calibration Performed At	-----

Detail of reference standards & Major equipments used

Equipment Name	RTD With Digital Indicator		
Make	YCT / SBI		
Model / SR No.	11080368		
Certificate No.	TL/021/219.1.1		
Calibration Validity	28/02/2022		
Calibration by	Tempsens		

Environmental Condition	Temperature	25±4 °C	Calibration Reference	IS:6274
	Relative Humidity	50±20 %	Calibration Procedure	CP-02T


Calibration Results

Serial No.	Std. Reading in (°C)	Measured Reading in (°C)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	2.11	2	±0.3 °C
02.	4.15	4	
03.	5.86	6	
04.	7.72	8	

Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor k = 2 such that the coverage probability corresponds to approximately 95%. (3) The above results are valid at the time of and under the stated conditions measurement. (4) This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By
(Calibration Engg./TM)
(D.B SAXSENA)
Form No. QF-47



Approved By
(QM/TM)
(NEERAJ TYAGI)
Page No. 1 of 1

---End of Report---