



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

Customer name And Address	M/S: Wellness Diagnostic Pvt. Ltd. Dharamveer Market, 2 nd Floor Near Seble Cinema Mathura Road Badarpur, New Delhi-110044.	ULR No.	CC223922000000763 F
		Service request No & date	P/02/21/02/2022
		Date of calibration	21/02/2022
		Valid up to	20/02/2023
		Date of issue	22/02/2022

Instrument Details

Instrument name	Micro Pipette	Sr. No.	----
Make / Model No.	Dragon Lab	Location	Lab
Range / Size	10 to 100 µl	Accuracy	-----
Least Count	1µl	Visual Inspection	OK
I.D. No.	P-01	Calibration Performed At	-----

Detail of reference standards & Major equipments used

Equipment Name	A set of weight box	Weighing Balance	
Make	WEIGHTRONICS	AND	
Model / SR No.	-----	GH-252/ 78002	
Certificate No.	TYCON/WB/12/274	TYE/WB/02/2022/022	
Calibration Validity	25/12/2023	20/02/2023	
Calibration by	TYCON ENGINEERING	Tycon Engineering	

Environmental Condition	Temperature	23± 3 °C	Calibration Reference	IS:1997
	Relative Humidity	50±20%	Calibration Procedure	CP-33

Calibration Results

Serial No.	Nominal Capacity in (µl)	Measured Capacity in (µl)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	20	19.96	±0.21 µl
02.	50	49.97	
03.	100	99.92	

Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor k = 2 such that the coverage probability corresponds to approximately 95%. (3) The above results are valid at the time of and under the stated conditions measurement. (4) This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

asian technology

Calibrated By
(Calibration Engg./TM)
(MAHIPAL)
Form No. QF-47



Approved By
(QM/TM)
(NEERAJ TYAGI)
Page No. 1 of 1

---End of Report---