


INSTALLATION QUALIFICATION (IQ)

INSTRUMENT DESCRIPTION

MODEL NAME : URIT 50
SERIAL NO. : 5020845E
SOFTWARE VERSION : 13.02.1
MANUFACTURER : URIT MEDICAL ELECTRONIC CO. LTD.
CUSTOMER NAME : OLIVE HEART & DIAGNOSTICS, MOHALI
DATE OF INSTALLATION : 06/10/2021
DATE OF CALIBRATION : 06/10/2021
NAME :
DESIGNATION :
COMPANY :
CUSTOMER (SIGNATURE) : 
NAME : DR. SUCHITA GUPTA
DESIGNATION : DIRECTOR

Using URIT 11G strips checked the sample Repeatability. Run same sample for 3 times.

TEST DATAS in

PARAMETERS	READING 1	READING 2	READING 3
LEU(Cell/uL)	0	0	0
KET(mmol/L)	0	0	0
NIT	0	0	0
URO	NORMAL	NORMAL	NORMAL
BIL(umol/L)	0	0	0
PRO(g/L)	0	0	0
GLU(mmol/L)	0	0	0
SG	1.030	1.030	1.030
BLD(cells/uL)	0	0	0
pH	4.5	4.5	4.5
Vc (mmol/L)	5.6	5.6	5.6

Tested By



Checked By

