

ABIRAMI KIDNEY CARE

Dr. THANGAVELU HOSPITAL

No. 582, Brough Road, Near Ravi Theater, Erode - 638 011.

Phone : 0424-2269495, Cell : 95007 94485

e-mail : abiramikidneycare@gmail.com,



28/02/2022

Dear Sir / Madam,


Greetings from Abirami Kidney Care (P) Ltd.,

This is to inform you that, we are using semi auto analyser (Model- ADAM-AG) as stand by only. Because we are having ROCHE COBAS 400 plus fully automated analyser and we are using the same. Also IQ, OQ, PQ records are not applicable to the semi auto analyser as stated by the company personnels.

So we are submitting the calibration records for vetting.

Thank you.

With warm regards



For Abirami Kidney Care (P) Ltd.

Erode.

TEST & CALIBRATION CERTIFICATE

UNIT DESCRIPTION:

SEMI AUTO ANALYZER

COMPANY:

SA

S/N:

AT0621016SMD

MODEL:

ADAM-AG

NAME OF CUSTOMER
& ADDRESS:

ABIRAMI KIDNEY CARE

NO: 580, 581, BROUGH ROAD,

NEAR RAVI THEATRE,

ERODE-638 011.

SEMI AUTO ANALYZER SERVICE: PREVENTIVE MAINTENANCE DONESERVICE: OKCALIBRATION: OKINSTRUMENT WORKING CONDITION: GOOD

CALIBRATION DATE:

25.03.2021


CALIBRATION VALIDITY:

24.03.2022

DATE: 25.03.2021

TEST BY: G. DINESH

APPROVED BY: QA.



NOTE: 1) Calibration Points & due date is given as per customer request 2) This certificate refer only to particular item submitted for certificate 3) The calibration results reported are valid at the time of and under the stated conditions of the measurements 4) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate.

CALIBRATION CERTIFICATE

DATE: 25.03.2021

NAME OF CUSTOMER:

ABIRAMI KIDNEY CARE
NO:580,581, BROUGH ROAD
NEAR RAVI THEATRE,
ERODE-638001

ADAM-AG CHEMISTRY SEMI AUTO ANALYZER PHOTOMETER BEARING.
S/NO:AT0621016SMD CALIBRATION ON: 25.03.2021

CALIBRATION VALUES:

LAMP AMPLITUDE


FILTERS	AMP	DELTA	RD.NO
340	3.6679	3.6679	1
405	4.7900	4.7900	1
510	5.0087	5.0087	1
545	5.6256	5.6256	1
560	6.2156	6.2156	1
600	6.2028	6.2028	1
635	5.1256	5.1256	1
670	6.2569	6.2569	1

<u>VALUES</u>	<u>CURRENT</u>	<u>NEW</u>
<u>SAMPLE VOLUME</u>	<u>462</u>	<u>449</u>

**PERISTALTIC PUMP
ADJUSTMENT TEST**

NEXT CALIBRATION DUE ON: 24.3.2022

TESTED BY: G DINESH

APPROVED BY: 

NOTE: 1) Calibration Points & due date is given as per customer request 2) This certificate refer only to particular test submitted for certificate 3) The calibration results reported are valid at the time of and under the stated conditions of the measurements 4) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate.