

### CALIBRATION CERTIFICATE

**NAME OF CUSTOMER:**  
**ABIRAMI KIDNEY CARE**  
**NO:580,581, BROUGH ROAD**  
**NEAR RAVI THEATRE,**  
**ERODE-638001**

**DATE: 25.03.2021**

**DESCRIPTION OF THE EQUIPMENT: CENTRIFUGE**

**COMPANY: REMI**

**MODEL: R-8C -(REMI)**

**RPM RANGE:1000-4500**

**SL.NO: ZFIN-31556**

**INSTRUMENT USED FOR CALIBRATION: DIGITAL TACHOMETER**

**MODEL NO:KUSAM-MECO&KM-2234BL**

**SL.NO: 06887**

**CERTIFIED BY: RELTEC CALIBRATION PVT.LTD**

| S.NO | INSTRUMENT SPEED | TACHOMETER READING | ALLOWABLE ERROR | ERROR OBSERVED ± | RESULT |
|------|------------------|--------------------|-----------------|------------------|--------|
| 1.   | 1000             | 1019               | ±100            | 19               | OK     |
| 2.   | 2500             | 2492               | ±100            | 8                | OK     |
| 3.   | 4000             | 3990               | ±100            | 10               | OK     |

**NEXT CALIBRATION DUE ON: 24.03.2022**

**TESTED BY: G DINESH**

**APPROVED BY: G DINESH**

NOTE: 1) Calibration Points & due date is given as per customer request 2) This certificate is valid only to particular item submitted for certificate 3) The calibration results reported are valid only at the time of and under the stated conditions of the measurements 4) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate.

