

SI. NO.	DISTRICT NAME	Unique ID	ICTC CENTRE
1	Alappuzha	KL-SRL-01-IC01	General Hospital Alappuzha
2	Alappuzha	KL-SRL-01-IC02	THQH Cherthala
3	Alappuzha	KL-SRL-01-IC03	Dept of Microbiology ,MCH,VANDANAM
4	Alappuzha	KL-SRL-01-IC04	THQH Harippad
5	Alappuzha	KL-SRL-01-IC05	THQH Kayamkulam
6	Alappuzha	KL-SRL-01-IC06	Dept of O&G, MCH,VANDANAM
7	Alappuzha	KL-SRL-01-IC07	District hospital , Mavelikara
8	Alappuzha	KL-SRL-01-IC08	Thaluk Hospital Thuravoor
9	Alappuzha	KL-SRL-01-IC09	W & C Hospital, Alappuzha
10	Alappuzha	KL-SRL-01-IC10	District hospital ,Chengannur
11	Ernakulam	KL-SRL-01-IC11	W&C, Mattanchery
12	Ernakulam	KL-SRL-01-IC12	General Hospital , Ernakulam
13	Ernakulam	KL-SRL-01-IC13	THQH, Thripunithura
14	Ernakulam	KL-SRL-01-IC14	THQH ,Muvatupuzha
15	Ernakulam	KL-SRL-01-IC15	THQH, Perumbavur
16	Ernakulam	KL-SRL-01-IC16	Thaluk Hospital ,Piravom
17	Ernakulam	KL-SRL-01-IC17	CHC, Kalady
18	Ernakulam	KL-SRL-01-IC18	CHC , Vengola
19	Ernakulam	KL-SRL-01-IC19	THQH, Fort Kohi
20	Ernakulam	KL-SRL-01-IC20	District Hospital, Aluva
21	Ernakulam	KL-SRL-01-IC21	THQH, North Paravoor
22	Ernakulam	KL-SRL-01-IC22	THQH , Kothamangalam
23	Ernakulam	KL-SRL-01-IC23	CHC, Angamali
24	Ernakulam	KL-SRL-01-IC24	Sub Jail, EKM
25	Ernakulam	KL-SRL-01-IC25	GoVT. Medical College, Ernakulam

Last year PT report (2021-22)

① EQAS - Proficiency Testing report Round I 2021-22

Report dispatch date - 27/09/22

② PT report Round II 2021-22

Report dispatch date - 21/2/2022



PC-1034

CMC VIROEQAS
By National Reference Laboratory (NRL)
Under National AIDS Control Organisation (NACO)
PT Unit, Department of Clinical Virology, Christian Medical College, Vellore-632004, Tamil Nadu
Email: viroeqas@cmcvellore.ac.in Phone: 0416-2283455

PANEL: HIV Serology

Laboratory ID: KL-SRL-01-IC01

Distribution Round: Round -2, 2021-2022

Panel dispatch date: 7-Dec-21

Result received month: 01-Jan-22

SL NO	SAMPLE ID	INTENDED RESULT	YOUR RESULT	MATCHING YES/NO	% MATCHING
1	S022101	Negative	Negative	YES	100
2	S022102	Positive for HIV-1	Positive for HIV-1	YES	100
3	S022103	Positive for HIV-1	Positive for HIV-1	YES	100
4	S022104	Negative	Negative	YES	100

REMARKS: PERFORMANCE IS SATISFACTORY

Comments:

Participants are scored based on qualitative result.

PT Unit, Department of Clinical Virology, Christian Medical College, Vellore
The data in this CMCVIROEQAS reports are confidential.

Page 1 of 2



PC-1034

CMC VIROEQAS
By National Reference Laboratory (NRL)
Under National AIDS Control Organisation (NACO)
PT Unit, Department of Clinical Virology, Christian Medical College, Vellore-632004, Tamil Nadu
Email: viroeqas@cmcvellore.ac.in Phone: 0416-2283455

Assigned Value:

Assigned value is determined using multiple assays/multiple testing of the same analyte. The proficiency testing material is a pooled lyophilized plasma sample prepared by following institutional proficiency testing standard operating protocol (PT-SOP) by trained staff. None of the work related to testing, preparation and packaging of sample is subcontracted. The homogeneity of the PT material is determined using multiple testing of multiple aliquots. The stability of the testing material is determined by assessing the reactivity of the specimen till the closing date.

Enquires: For queries, please contact CMCVIREQAS coordinator at email: nrl.cmc@gmail.com

Name of CMCVIREQAS coordinator

Signature

Dr. Rajesh Kannangai

Department of Clinical Virology

Report Dispatch Date: 21-Feb-22

Report authorized by: CMCVIREQAS Coordinator

END OF REPORT

PT Unit, Department of Clinical Virology, Christian Medical College, Vellore
The data in this CMCVIREQAS reports are confidential.



PC-1034

CMC VIROEQAS
By National Reference Laboratory (NRL)
Under National AIDS Control Organisation (NACO)
PT Unit, Department of Clinical Virology, Christian Medical College, Vellore-632004, Tamil Nadu
Email: viroeqas@cmcvellore.ac.in Phone: 0416-2283455

PANEL: HIV Serology

Laboratory ID: KL-SRL-01-IC01

Distribution Round: Round -1, 2021-2022

Panel dispatch date: 23/06/2021

Result received date: 30/07/2021


SL NO	SAMPLE ID	INTENDED RESULT	YOUR RESULT	MATCHING YES / NO	% MATCHING
1	S012101	Positive for HIV-1	Positive for HIV-1	YES	100
2	S012102	Negative	Negative	YES	100
3	S012103	Negative	Negative	YES	100
4	S012104	Positive for HIV-1	Positive for HIV-1	YES	100

REMARKS: PERFORMANCE IS SATISFACTORY

Comments:

Participants are scored based on qualitative result.

PT Unit, Department of Clinical Virology, Christian Medical College, Vellore
The data in this CMCVIROEQAS reports are confidential.



PC-1034

CMC VIROEQAS
By National Reference Laboratory (NRL)
Under National AIDS Control Organisation (NACO)
PT Unit, Department of Clinical Virology, Christian Medical College, Vellore-632004, Tamil Nadu
Email: viroeqas@cmcvellore.ac.in Phone: 0416-2283455

Assigned Value:

Assigned value is determined using multiple assays/multiple testing of the same analyte. The proficiency testing material is a pooled lyophilized plasma sample prepared by following institutional proficiency testing standard operating protocol (PT-SOP) by trained staff. None of the work related to testing, preparation and packaging of sample is subcontracted. The homogeneity of the PT material is determined using multiple testing of multiple aliquots. The stability of the testing material is determined by assessing the reactivity of the specimen till the closing date.

Enquires: For queries, please contact CMCVIROEQAS coordinator at email: nrl.cmc@gmail.com

Name of CMCVIROEQAS coordinator

Signature

Dr. Rajesh Kannangai

Department of Clinical Virology

Report Dispatch Date: 27/09/2021

Report authorized by: CMCVIROEQAS Coordinator

END OF REPORT

PT Unit, Department of Clinical Virology, Christian Medical College, Vellore
The data in this CMCVIROEQAS reports are confidential.