

VIEW LAB MONTHLY SUMMARY

Lab Name **POOJA DIAGNOSTICS**

[Details About Robust Analysis](#)

Lab No **2461**

[Detail About Monthly Summary](#)

Month **January**

[Detail about SDI](#)

Year **2022**

Constituent Group **Chemistry I**

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/01/2022

Date of Report Published : 07/02/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Agape	137	136.51	5.84	7.97	139 mg/dl	0.31	1.36
2	UREA	Urease UV / GLDH	Agape	134	48.51	7.80	3.78	47 mg/dl	-0.40	0.65
3	CREATININE	Enzymatic Colorimetric	Agape	50	1.64	7.18	0.12	3 mg/dl	11.33	0.03
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Agape	116	2.41	10.03	0.24	2.4 mg/dl	-0.04	0.04
5	T-PROTEIN	Biuret - Colorimetric	Agape	137	5.12	6.09	0.31	5.5 g/dl	1.22	0.05
6	ALBUMIN	BCG - colorimetric	Agape	131	3.43	6.35	0.22	3 g/dl	-1.97	0.04
7	CALCIUM	OCPC (O - Cresolphthalein Compleazone)	Any Analyser (Automation / Semi Automation)	166	9.18	8.08	0.74	10.2 mg/dl	1.37	0.12
8	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Agape	72	5.36	6.19	0.33	6.5 mg/dl	3.43	0.08
9	URIC ACID	Enzymatic / Uricase Colorimetric	Agape	128	7.14	9.15	0.65	9.1 mg/dl	3.00	0.12
10	CHOLESTEROL	CHOD-PAP	Agape	140	98.01	6.90	6.77	99 mg/dl	0.15	1.14
11	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Agape	130	129.27	6.78	8.76	136 mg/dl	0.77	1.54
12	HDL CHO	Direct method / Enzymatic colorimetric	Agape	107	24.20	11.77	2.85	27 mg/dl	0.98	0.55
13	SODIUM	ISE - Direct	Transasia / Erba	184	135.58	3.02	4.09	138.7 mmol/L	0.76	0.60
14	POTASSIUM	ISE - Direct	Transasia / Erba	184	3.64	4.56	0.17	3.8 mmol/L	0.96	0.02
15	CHLORIDE	ISE - Direct	Transasia / Erba	129	106.22	3.45	3.66	110 mmol/L	1.03	0.64
16	AST	UV kinetic(with & without PLP (P-5-P))	Agape	116	46.94	12.48	5.86	48 U/L	0.18	1.09
17	ALT	UV kinetic(with & without PLP (P-5-P))	Agape	105	117.50	17.76	20.87	115 U/L	-0.12	4.07
18	ALP	PNP AMP kinetic	Agape	42	78.72	12.21	9.61	88 U/L	0.97	2.97
19	AMYLASE	CNPG3	Agape	54	36.24	14.56	5.28	38 U/L	0.33	1.44

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out