

VIEW LAB MONTHLY SUMMARY

Lab Name **POOJA DIAGNOSTICS**

[Details About Robust Analysis](#)

Lab No **2461**

[Detail About Monthly Summary](#)

Month **December**

[Detail about SDI](#)

Year **2021**

Constituent Group **HbA1c**

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 18/12/2021

Date of Report Published : 03/01/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	131	6.76	10.61	0.72	6.8 %	0.06	0.13

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out