

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE



Customer Name & Address:	M/S. New Green Lab& Biopsy Centre, #1/786, Bharathi Nagar 7 th Street, Ramanathapuram – 623 503	Certificate No:	AMS/HOSP/289/0523
		Issuing Date:	16/02/2022
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Digital Thermometer	Id. No:	NGC/LAB /DT /001
Make:	Symbol	Department:	Laboratory
Model:	Lab	Calibrated On:	15/02/2022
Sl. No:	---	Calibration Due Suggested:	14/02/2022
Type:	Digital PT 100	DUC Condition on Receipt:	OK
Range & Least Count:	-15 to +15 °C / LC 0.1°C		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	PC/01/22/ATL/0171-01	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

Test Results:

Sl. No	Freezer Readings (°C)			Chiller (°C)			Chamber (°C)		
	DUC	STD	Correction	DUC	STD	Correction	DUC	STD	Correction
1	-2.0	-1.84	+0.16	1.0	1.09	+0.09	2.0	2.08	+0.08
2	-2.5	-2.38	+0.12	2.0	2.12	+0.12	2.5	2.61	+0.11
3	-3.0	-2.85	+0.15	2.5	2.65	+0.15	3.0	3.12	+0.12
4	-3.5	-3.33	+0.17	3.0	3.16	+0.16	3.5	3.62	+0.12
5	-4.0	-3.83	+0.17	3.5	3.66	+0.16	4.0	4.13	+0.13
6	-4.5	-4.39	+0.11	4.0	4.18	+0.18	4.5	4.61	+0.11
7	-5.0	-4.92	+0.08	4.5	4.68	+0.18	5.0	5.12	+0.12
8	-5.5	-5.45	+0.05	5.0	5.19	+0.19	5.5	5.62	+0.12
9	-6.0	-6.04	-0.04	5.5	5.69	+0.19	6.0	6.11	+0.11
10	-6.5	-6.48	+0.02	6.0	6.23	+0.23	6.5	6.63	+0.13
11	-7.0	-7.05	-0.05	6.5	6.71	+0.21	7.0	7.14	+0.14
12	-7.5	-4.56	-0.06	7.0	7.22	+0.22	7.5	7.64	+0.14
13	-8.0	-8.06	-0.06	7.5	7.73	+0.23	8.0	8.12	+0.12
14	-8.5	-8.54	-0.04	8.0	8.25	+0.25	8.5	8.61	+0.11

Allowable Tolerance: User Defined.

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)

B. Jeyaraj

© aceagm@gmail.com



©www.acemedical.in

Approved by:
(Technical Manager)

B. Jeyaraj

93455 56140, 93455 56148