



Cert No. CC-2543



Medical Engineering & Services

Things Beyond Thoughts

(An ISO/IEC 17025:2005 NABL Accredited Calibration Laboratory)

Abhcon Crown, X1/411/21, Ambadi Lane
Kokkalai, Thrissur-21. Ph: 0487 2420367
Email: mescalibration.tsr@gmail.com
Visit us @www.mescalibration.com

CALIBRATION CERTIFICATE

Certificate Number : MES/220921/C3 ULR Number : CC254321000000702F
Customer Details : NATIONAL AYURVEDA RESEARCH INSTITUTE FOR PANCHAKARMA, CHERUTHURUTHY.
Date of Calibration : 22/09/2021 Date of SRF : 22/09/2021
Next Calibration Due : 21/09/2022 Date of Issue : 22/09/2021

DETAILS OF DEVICE UNDER CALIBRATION(DUC)

Nomenclature : MICROPIPETTE Serial No. : V34735
Make / Model : ACCUPIPET ID No. : NA
Equipment Condition : GOOD Resolution : 0.2µL
Range : 20-200µL Location : BIO CHEMISTRY

DETAILS OF MASTER EQUIPMENT USED

Nomenclature : Electronic Precision Balance Range : 0 – 220 gm
Make : Boeco Readability : 0.01/0.1 mg
Model : BBC-22 Serial No. : 27104178
Certificate No. : MES/17012020/EB/C1 Calibration Due Date : 13.01.2022

ENVIRONMENTAL CONDITIONS

Temperature(°C)	Relative Humidity (%)	Atmospheric Pressure (mbar)	Water Temperature(°C)
27.3	55	1008	27.6

CALIBRATION RESULTS

Discipline		Mechanical		Parameter		Volume	
Sl. No.	Set Volume	Standard Value	Systematic Error	Acceptable Systematic Error	Random Error	Acceptable Random Error	Expanded Uncertainty
	µl	µl	µl	± µl	µl	± µl	± µl
1	20	20.03	-0.03	0.20	0.11	0.10	0.319
2	100	100.36	-0.36	0.80	0.10	0.30	0.301
3	200	199.85	0.15	1.60	0.51	0.60	1.492

Notes:-

- The results in this certificate are only related to DUC submitted calibration and valid at the time of calibration under stated condition of measurements
- The measurement uncertainty is estimated at a confidence level of 95.45% with coverage factor k=2.
- Results in this certificate shall not be reproduced except in full without the written approval of this centre.
- Next calibration due date is mentioned as per customer requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Acceptable error limit is given as per ISO 8655-2:2002(E) Guidelines
- Calibration is carried out as per ISO 8655-6 Guidelines; MES work procedure: MES/WP/VOLUME.
- The tests and /or Calibration marked with an * are not accredited by NABLs

Calibrated by
Prabin KP
(Calibration Engineer)



Authorized Signatory
Binoj Thomas
(Technical Manager)