



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - November 2021



Lab Name	POOJA DIAGNOSTICS	Lab No	2461
Constituent Group	HbA1c	Date of Result Entered :	16/11/2021
PT item	Lyophilized Whole blood	Date of Report Published :	03/12/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	143	7.88	12.33	0.97	8.7 %	0.84	0.16
SDI Range					Interpretation					
Within -1.00 to +1.00					Excellent.					
Within ±1.01 to ±2.00					Good.					
Within ±2.01 to ±2.99					Accept with caution. Warning Signal.					
Beyond ±3.0					Unacceptable performance. Action Signal.					

LAB ADDRESS :

POOJA DIAGNOSTICS
NO.8 ABOVE CANARA BANK, 80 FEET ROAD, CHANDRA LAYOUT
BENGALURU
KARNATAKA560040



Coordinator Contact Details:	Dr. Pamela Christudoss
Email:clinqc@cmcvellore.ac.in	CMC EQAS Coordinator
Contact Number: 0416-2283102	Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - November 2021



Lab Name	POOJA DIAGNOSTICS	Lab No	2461
Constituent Group	Chemistry II	Date of Result Entered :	17/11/2021
PT item	Lyophilized Serum	Date of Report Published :	03/12/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Agape	291	146.01	8.71	12.71	137 mg/dl	-0.71	1.49
2	UREA II	Urease UV / GLDH II	Agape	233	42.58	11.84	5.04	42 mg/dl	-0.12	0.66
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Agape	210	8.04	11.57	0.93	9.8 mg/dl	1.89	0.13
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Agape	223	2.38	15.69	0.37	1.9 mg/dl	-1.28	0.05
5	T-PROTEIN II	Biuret - Colorimetric II	Agape	175	5.32	7.86	0.42	4.7 g/dl	-1.48	0.06
6	ALBUMIN II	BCG - Colorimetric II	Agape	169	3.27	9.69	0.32	2.95 g/dl	-1.01	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Agape	210	7.93	11.14	0.88	8.7 mg/dl	0.87	0.12
8	CHOLESTROL II	CHOD-PAP II	Agape	279	106.59	11.53	12.29	91 mg/dl	-1.27	1.47
9	TRIGLYCERID E II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	196	153.58	10.53	16.18	137 mg/dl	-1.02	2.31

SDI Range	Interpretation
-----------	----------------

Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

POOJA DIAGNOSTICS

NO.8 ABOVE CANARA BANK, 80 FEET ROAD, CHANDRA LAYOUT

BENGALURU

KARNATAKA560040



Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College,
Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - October 2021



Lab Name	POOJA DIAGNOSTICS	Lab No	2461
Constituent Group	HbA1c	Date of Result Entered :	16/10/2021
PT item	Lyophilized Whole blood	Date of Report Published :	01/11/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	145	8.96	15.36	1.38	9.4 %	0.32	0.23
SDI Range					Interpretation					
Within -1.00 to +1.00					Excellent.					
Within ±1.01 to ±2.00					Good.					
Within ±2.01 to ±2.99					Accept with caution. Warning Signal.					
Beyond ±3.0					Unacceptable performance. Action Signal.					

LAB ADDRESS :

POOJA DIAGNOSTICS
NO.8 ABOVE CANARA BANK, 80 FEET ROAD, CHANDRA LAYOUT
BENGALURU
KARNATAKA560040



Coordinator Contact Details:	Dr. Pamela Christudoss
Email:clinqc@cmcvellore.ac.in	CMC EQAS Coordinator
Contact Number: 0416-2283102	Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - October 2020



Lab Name	POOJA DIAGNOSTICS	Lab No	2461
Constituent Group	Chemistry II	Date of Result Entered :	17/10/2020
PT item	Lyophilized Serum	Date of Report Published :	09/11/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Mindray	132	171.06	9.54	16.32	175 mg/dl	0.24	2.84
2	UREA I	Urease UV / GLDH	Mindray	117	109.77	13.39	14.70	106 mg/dl	-0.26	2.72
3	CREATININE I	Jaffes Kinetic - Alkaline picrate	Mindray	89	4.62	13.35	0.62	4.1 mg/dl	-0.84	0.13
4	T.BILIRUBIN I	OTHERS (Any Other Principles / Methods)	Any Analyser	153	3.56	16.77	0.60	3.5 mg/dl	-0.10	0.10
5	T-PROTEIN I	Biuret - Colorimetric	Mindray	100	5.00	11.76	0.59	5.4 g/dl	0.68	0.12
6	ALBUMIN I	BCG - colorimetric	Mindray	101	3.05	8.82	0.27	2.9 g/dl	-0.56	0.05
7	URIC ACID I	Enzymatic / Uricase Colorimetric	Mindray	91	4.26	19.82	0.84	4.2 mg/dl	-0.07	0.18
8	CHOLESTEROL I	CHOD-PAP	Mindray	117	99.97	10.01	10.01	98 mg/dl	-0.20	1.85
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Mindray	108	189.54	12.58	23.84	188 mg/dl	-0.06	4.59

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
POOJA DIAGNOSTICS
NO.8 ABOVE CANARA BANK, 80 FEET ROAD, CHANDRA LAYOUT
BENGALURU
KARNATAKA560040



Coordinator Contact Details:	Dr. Pamela Christudoss
Email:clinqc@cmcvellore.ac.in	CMC EQAS Coordinator
Contact Number: 0416-2283102	Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - September 2021



Lab Name	POOJA DIAGNOSTICS	Lab No	2461
Constituent Group	HbA1c	Date of Result Entered :	16/09/2021
PT item	Lyophilized Whole blood	Date of Report Published :	05/10/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	156	7.14	11.83	0.84	6.9 %	-0.28	0.14
SDI Range					Interpretation					
Within -1.00 to +1.00					Excellent.					
Within ±1.01 to ±2.00					Good.					
Within ±2.01 to ±2.99					Accept with caution. Warning Signal.					
Beyond ±3.0					Unacceptable performance. Action Signal.					

LAB ADDRESS :

POOJA DIAGNOSTICS
NO.8 ABOVE CANARA BANK, 80 FEET ROAD, CHANDRA LAYOUT
BENGALURU
KARNATAKA560040



Coordinator Contact Details:	Dr. Pamela Christudoss
Email:clinqc@cmcvellore.ac.in	CMC EQAS Coordinator
Contact Number: 0416-2283102	Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - September 2021



Lab Name	POOJA DIAGNOSTICS	Lab No	2461
Constituent Group	Chemistry II	Date of Result Entered :	16/09/2021
PT item	Lyophilized Serum	Date of Report Published :	05/10/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Agape	285	403.14	9.49	38.24	409 mg/dl	0.15	4.53
2	UREA II	Urease UV / GLDH II	Agape	233	91.04	10.17	9.26	89 mg/dl	-0.22	1.21
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Agape	206	3.00	9.93	0.30	3.3 mg/dl	1.01	0.04
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Agape	216	2.21	15.29	0.34	2.2 mg/dl	-0.03	0.05
5	T-PROTEIN II	Biuret - Colorimetric II	Agape	169	5.17	8.70	0.45	5.1 g/dl	-0.16	0.07
6	ALBUMIN II	BCG - Colorimetric II	Agape	163	3.24	8.83	0.29	3.1 g/dl	-0.49	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Agape	215	8.28	10.25	0.85	9.9 mg/dl	1.91	0.12
8	CHOLESTROL II	CHOD-PAP II	Agape	278	108.09	11.39	12.32	103 mg/dl	-0.41	1.48
9	TRIGLYCERID E II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	191	224.03	10.64	23.84	224 mg/dl	0.00	3.45

SDI Range	Interpretation
-----------	----------------

Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

POOJA DIAGNOSTICS

NO.8 ABOVE CANARA BANK, 80 FEET ROAD, CHANDRA LAYOUT

BENGALURU

KARNATAKA560040



Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College,
Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - August 2021



Lab Name	POOJA DIAGNOSTICS	Lab No	2461
Constituent Group	HbA1c	Date of Result Entered :	16/08/2021
PT item	Lyophilized Whole blood	Date of Report Published :	01/09/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	131	7.61	10.53	0.80	7.9 %	0.36	0.14
SDI Range					Interpretation					
Within -1.00 to +1.00					Excellent.					
Within ±1.01 to ±2.00					Good.					
Within ±2.01 to ±2.99					Accept with caution. Warning Signal.					
Beyond ±3.0					Unacceptable performance. Action Signal.					

LAB ADDRESS :

POOJA DIAGNOSTICS
NO.8 ABOVE CANARA BANK, 80 FEET ROAD, CHANDRA LAYOUT
BENGALURU
KARNATAKA560040



Coordinator Contact Details:	Dr. Pamela Christudoss
Email:clinqc@cmcvellore.ac.in	CMC EQAS Coordinator
Contact Number: 0416-2283102	Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME



MONTHLY SUMMARY REPORT - August 2021

Lab Name	POOJA DIAGNOSTICS	Lab No	2461
Constituent Group	Chemistry II	Date of Result Entered :	16/08/2021
			1
PT item	Lyophilized Serum	Date of Report Published :	01/09/2021
			1

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Agape	239	258.66	8.15	21.08	258 mg/dl	-0.03	2.73
2	UREA II	Urease UV / GLDH II	Agape	215	43.28	13.10	5.67	42 mg/dl	-0.23	0.77
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Agape	180	1.73	11.04	0.19	1.6 mg/dl	-0.68	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Agape	176	4.16	13.50	0.56	4.1 mg/dl	-0.11	0.08
5	T-PROTEIN II	Biuret - Colorimetric II	Agape	144	5.31	9.08	0.48	5.4 g/dl	0.19	0.08
6	ALBUMIN II	BCG - Colorimetric II	Agape	136	3.31	9.07	0.30	3.4 g/dl	0.30	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Agape	187	3.20	33.71	1.08	4.9 mg/dl	1.58	0.16
8	CHOLESTROL II	CHOD-PAP II	Agape	245	106.11	12.35	13.11	115 mg/dl	0.68	1.67
9	TRIGLYCERID E II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	160	125.10	12.57	15.73	125 mg/dl	-0.01	2.49

SDI Range	Interpretation
-----------	----------------

Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

POOJA DIAGNOSTICS

NO.8 ABOVE CANARA BANK, 80 FEET ROAD, CHANDRA LAYOUT

BENGALURU

KARNATAKA560040



Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College,
Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****

BENGALURU
KARNATAKA560040

Coordinator Contact Details:
Email:clinqc@cmevellore.ac.in
Contact number:0416-2283102



Dr. Pamela Christudoss
CMC EQAS
Coordinator
Christian Medical
College, Vellore

Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

12	HDL CHO	Direct method / Enzymatic colorimetric	Agape	107	24.20	11.7 7	2.85	27 mg/dl	0.98	0.5 5
13	SODIUM	ISE - Direct	Transasia / Erba	184	135.5 8	3.02	4.09	138.7 mmol/ L	0.76	0.6 0
14	POTASSIUM	ISE - Direct	Transasia / Erba	184	3.64	4.56	0.17	3.8 mmol/ L	0.96	0.0 2
15	CHLORIDE	ISE - Direct	Transasia / Erba	129	106.2 2	3.45	3.66	110 mmol/ L	1.03	0.6 4
16	AST	UV kinetic(with & without PLP (P-5-P))	Agape	116	46.94	12.4 8	5.86	48 U/L	0.18	1.0 9
17	ALT	UV kinetic(with & without PLP (P-5-P))	Agape	105	117.5 0	17.7 6	20.8 7	115 U/L	- 0.12	4.0 7
18	ALP	PNP AMP kinetic	Agape	42	78.72	12.2 1	9.61	88 U/L	0.97	2.9 7
19	AMYLASE	CNPG3	Agape	54	36.24	14.5 6	5.28	38 U/L	0.33	1.4 4

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

POOJA DIAGNOSTICS

NO.8 ABOVE CANARA BANK, 80 FEET ROAD, CHANDRA LAYOUT

BENGALURU

KARNATAKA560040



Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-
2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College,
Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - December 2021



Lab Name	POOJA DIAGNOSTICS	Lab No	2461
Constituent Group	HbA1c	Date of Result Entered :	18/12/2021
PT item	Lyophilized Whole blood	Date of Report Published :	03/01/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	131	6.76	10.61	0.72	6.8 %	0.06	0.13
SDI Range					Interpretation					
Within -1.00 to +1.00					Excellent.					
Within ±1.01 to ±2.00					Good.					
Within ±2.01 to ±2.99					Accept with caution. Warning Signal.					
Beyond ±3.0					Unacceptable performance. Action Signal.					

LAB ADDRESS :

POOJA DIAGNOSTICS
NO.8 ABOVE CANARA BANK, 80 FEET ROAD, CHANDRA LAYOUT
BENGALURU
KARNATAKA560040



Coordinator Contact Details:	Dr. Pamela Christudoss
Email:clinqc@cmcvellore.ac.in	CMC EQAS Coordinator
Contact Number: 0416-2283102	Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - December 2021



Lab Name	POOJA DIAGNOSTICS	Lab No	2461
Constituent Group	Chemistry II	Date of Result Entered :	18/12/2021
PT item	Lyophilized Serum	Date of Report Published :	03/01/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Agape	280	221.94	8.18	18.16	219 mg/dl	-0.16	2.17
2	UREA II	Urease UV / GLDH II	Agape	213	110.83	10.44	11.58	110.4 mg/dl	-0.04	1.59
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Agape	216	2.03	8.47	0.17	2 mg/dl	-0.17	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Agape	213	3.09	14.16	0.44	2.9 mg/dl	-0.43	0.06
5	T-PROTEIN II	Biuret - Colorimetric II	Agape	184	5.14	9.25	0.48	4.9 g/dl	-0.51	0.07
6	ALBUMIN II	BCG - Colorimetric II	Agape	161	3.22	10.46	0.34	3.2 g/dl	-0.06	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Agape	206	3.79	18.99	0.72	4.9 mg/dl	1.54	0.10
8	CHOLESTROL II	CHOD-PAP II	Agape	281	103.44	10.38	10.73	93 mg/dl	-0.97	1.28
9	TRIGLYCERID E II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	186	90.90	13.77	12.52	89 mg/dl	-0.15	1.84

SDI Range	Interpretation
-----------	----------------

Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

POOJA DIAGNOSTICS

NO.8 ABOVE CANARA BANK, 80 FEET ROAD, CHANDRA LAYOUT

BENGALURU

KARNATAKA560040



Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College,
Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****

QC Data Summary

Control: ERBA PATH

Start Date: 2/8/2022

End Date: 2/8/2022

Test	Results	Units	Time	Control Range	Remarks/Flag
PHOS	8.0	mg/dL	2/8/2022 2:55:47 PM	7.560-9.040	
CALC	10.9	mg/dL	2/8/2022 11:42:19 AM	9.540-11.660	
ALB	5.15	g/dL	2/8/2022 11:42:19 AM	4.5200-5.5200	
PROT	9.28	g/dL	2/8/2022 11:42:19 AM	8.1800-10.0200	
ALP	522	U/L	2/8/2022 11:42:19 AM	428.96-561.04	
SGPT	129	U/L	2/8/2022 11:42:19 AM	124.78-163.22	
BILT	4.9	mg/dL	2/8/2022 11:42:19 AM	3.860-5.420	
URIC	12.7	mg/dL	2/8/2022 11:42:19 AM	10.980-13.420	
CREAT	3.58	mg/dL	2/8/2022 11:42:19 AM	2.8700-3.7900	
UREA	98.3	mg/dL	2/8/2022 11:42:19 AM	89.020-108.780	
TRIG	157	mg/dL	2/8/2022 11:42:19 AM	152.96-187.04	
CHOL	257	mg/dL	2/8/2022 11:42:19 AM	237.58-290.42	
GLU	245	mg/dL	2/8/2022 11:42:19 AM	221.40-270.60	

History Calibration Results

Test: CREAT

Print Date: 2/8/2022 6:25:09 PM

Request Date	Rule	Status	Reagent Blank	SD	
K	R0	A	B	C	D
2/8/2022 11:22:33 AM	Two-point Linear	Finished	-186.62578	0	
		151.59018	0.00000		

History Calibration Results

Test: PHOS

Print Date: 2/8/2022 6:27:32 PM

Request Date	Rule	Status	Reagent Blank	SD	
K	R0	A	B	C	D
2/8/2022 12:10:02 PM	Two-point Linear	Finished	2643.33023	0	
		668.63509	96.11375		

History Calibration Results

Test: URIC

Print Date: 2/8/2022 6:29:28 PM

Request Date	Rule	Status	Reagent Blank	SD	
K	R0	A	B	C	D
1/18/2022 10:59:08 AM	Two-point Linear	Finished	-254.79987	0	
		146.52367	-9.26211		
