

# CERTIFICATE OF CALIBRATION

Customer Name & Address	District Hospital Female Hardoi Uttar Pradesh 241001 India	Certificate No.	Z2LHDWH64
		Date of Calibration	17.02.2022
		Valid Upto	16.02.2023
		Date of Issue	18.02.2022

Instruments Details			
Instrument Name	CENTRIFUGE	S No.	na
Make/Model No.	Local Mnf./na	Location	ICTC DEPT.
ID No.	MHUP - 371-0036	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used			
Equipment Name	ELECTRICAL SAFETY ANALYZER	DIGITAL TACHO METER	THERMO HYGROMETER
Make	RIGEL MEDICAL	METRAVI	HTC-2
Model/SR No.	SAFE TEST 60/49N-0257	NCTM 1000/170603632	NA
Calibration Validity	13/12/2022	20/07/2022	20/07/2022
Calibration By	SEAWARD ELECTRONIC	NENO TECHNICAL SERVICES	UNIQUE CALIBRATION SOLUTIONS

Environmental Condition			
Temperature	27 °C	Relative Humidity	56 %

Calibration Results				
SL NO	SET READING (RPM)	MEASURED READING (RPM)	TOLERANCE(°C)	STATUS
1	1000	1008	±3%	P
2	2000	1981	±3%	P
3	2500	2494	±3%	P

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By  
(SE)



Approved By  
(QM)



**Cyrix Healthcare Pvt. Ltd.**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph.: +91 484 6544033, Mob: 9745555073  
Email : sales@cyrixhealthcare.com  
Service@cyrixhealthcare.com

**CYRIX**  
HEALTH CARE PVT LTD

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)