

ELECTRICAL SAFETY TEST VALUES

Certificate No.:-Z2LHDWH64
Equipment Name:-Centrifuge

BM ID:-MHUP - 371-0036
Model/Make:- Local Mnf./na

TEST NO.	TEST	READING VALUE					READING UNIT	HIGH LIMITS	LOW LIMITS	STATUS
1	Main Voltage: Live to Neutral	230.3					V	234.00	226.00	P
2	Main Voltage: Live to Earth	228.6					V	233.00	227.00	P
3	Main Voltage: Neutral to Earth	1.3					V	3.00	0.00	P
4	Equipment Current	0.3					A	16.00	0.00	P
5	Protective Earth Resistance	0.3					Ω	2.00	0.00	P
6	Earth Leakage Current-Normal Condition	204					μA	250.00	50.00	P
7	Earth Leakage Current-Open Condition	302					μA	500.00	250.00	P
8	Enclosure Leakage Current-Normal Condition	1.5					μA	10.00	0.00	P
9	Enclosure Leakage Current-Open Condition	0.8					μA	10.00	0.00	P
10	Patient Auxiliary-Normal Condition	NA	NA	NA	NA	NA	MΩ	1.00	0.00	P
11	Patient Auxiliary-Open Condition	NA	NA	NA	NA	NA	MΩ	1.00	0.00	P
12	Insulation Resistance: Mains-Protective Earth	NA					MΩ	2.00	0.00	P
13	Insulation Resistance: Applied Part-Protective Earth	NA					MΩ	2.00	0.00	P
14	Insulation Resistance: Mains-Applied Part	OL					MΩ	5.00	0.00	P
15	Insulation Resistance: Mains-NE	OL					MΩ	2.00	0.00	P
16	Insulation Resistance: Applied Part-NE	OL					MΩ	2.00	0.00	P

Service Engineer (SE)



Quality Manager (QM)

