



ADARSHA CALIBRATION SERVICES

NABL ACCREDITED CALIBRATION LABORATORY

Plot No.18,19, Survey No-118, Byraveshwara Nagar, Manganahalli Cross, Ullal Main Road,
Ullal (P), Bangalore-560056, Ph: +91-80-2324 0578, Mob: +91-9945086171/172/173

E: acscaLib@gmail.com / enquiry@adarshacalibrations.com W: www.adarshacalibrations.com



CERTIFICATE OF CALIBRATION

Page 1 of 1

Certificate Number: ACS/MC/20-21/SOI773-01 ULR Number: ULR-CC-272220000003439F

Customer Name & Address: M/s.Dr. Raju's Blooms Diagnostics Centre,
GangammaGudi Road, Beside Lakkamma Doddamuniyappa
Kalyanamaitappa, Chikkaballap 'I'-562101.

Date Of SRF: 19-08-2021 Calibrated At: Site
Date of Calibration: 19-08-2021 Calibration WINo: WI/MC/10
Next Cal Due: 18-08-2022 Date of Issue: 21-08-2021
Environmental Conditions: Temperature: 24.5°C Humidity: 53% R.H

DETAILS OF DEVICE UNDER CALIBRATION

Nomenclature: RPM Meter (Centrifuge) Serial. No: ZBEN-18505
Make & Model: Remi & R-8C Code/ID No: NA
Range: 0 to 5600 RPM Resolution: 10 RPM
Location: Biochemistry

DETAILS OF STANDARD INSTRUMENT USED

Nomenclature: Digital Tachometer Serial. No: AE.75584
Make / Model: Lutron & DT-1236L Cal Due Date: 13-03-2022
Range: 10to 99999RPM Certificate No: ACS/MC/20-21/102284-,01

CALIBRATION RESULTS:

Parameter: SPEED Discipline : Mechanical

Sl.No.	Standard Reading (RPM)	DUC Reading (RPM)	Error Obtained -(RPM)	Expanded Measurement Uncertainty +(RPM)
PHOTO				
1	1001			
2	2503			
3	4005			
4	5607			

REMARKS & CONCLUSION:

- The Measurement Uncertainty is estimated at a confidence level of 95.45 % with a coverage factor k=2.0.
- Next Calibration Due Date mentioned as per customer requirement.

Calibrated By

Vajresh N
Calibration Engineer



Authorized Signatory

Manjunath I
Technical Manager