


MEASURE TECHNO LAB 2, B.T. ROAD (JAYANTI CINEMA COMPLEX) BARRACKPORE, KOLKATA - 700120, W.B. Phone : 033 - 2215 - 0032, 2215 - 9687, 8100875519, Mobile: 9831190974 , LAB:- 8100143376, E-mail: measuretechno@yahoo.co.in	Form No. - MTL/22/2006	 <small>NABL ACCREDITED LABORATORY Certificate No. CC-3515</small>
	CALIBRATION CERTIFICATE OF AUTO PIPETTE	
CALIBRATION CERTIFICATE NO.: MTL / HOP(LN-II) / R02 / 09 - 21		ULR - CC254521000018554F
Page: 1 of 1		

1.0 Service Request No.: MTL / 14E / 09 / 21 - 22

1.1 Issued to: House of Pathology Labs Pvt. Ltd.
L - 113, Lajpat Nagar - 2,
New Delhi - 24.

1.2 Description & Identification of item to be Calibrated:	a) Name: Auto Pipette	b) Code No.: N.A.
	c) Sl. No.: YE20AAS0047231	d) Make: Dragon Lab
	e) Model / Type: N.A.	f) Range: 5 µl to 50 µl
	g) Sensor: N.A.	h) Resolution: 0.5 µl
	i) End User: Lab	j) Accuracy: ± 0.50 µl
	k) Calibration done at: On Site / √ In House	

1.3 Date of receipt of item : 09-09-21

1.4 Physical Condition of DUC : OK

1.5 Date of calibration : 14-09-21

1.6 Recommended date of next calibration : 14-09-22

1.7 Date of Issue : 16-09-21

1.8 Environmental Conditions During Calibration:

Temperature:	23 °C ± 2 °C
Humidity:	50 % RH ± 10 % RH
Pressure:	1005.6 mbar

1.9 Method of Calibration: SOP / MASS / 02 (As Per ISO : 8655 – 6 : 2002)

2.0 Traceability :

a) Standards used for calibration are traceable to National standards through NABL Accredited Laboratory.

b) The following standards / Equipment have been used.

- i) Weight Box (E1) Cal. Certificate No. WI / May / 19 / 010 (WEIGH INDIA, New Delhi) (Cal. Date: 17/05/19, Due Date: 17/05/22)
- ii) Data Acquisition Switch Unit Cal. Certificate No. JRPM - CCTR - ET - 2021 - 1227 (JRPM, Chennai) (Cal. Date: 26/08/21, Due Date: 25/08/22)
- iii) RTD (PT - 100) Cal. Certificate No. TL / 020 / 891.2.1 (TEMPSSENS, Udaipur) (Cal. Date: 27/10/20, Due Date: 26/10/21)

2.1 Result :

Mechanical Calibration

Sl. No.	Parameter/ Range	Nominal Value µl	Mass of Water mg	Volume of Water at 20 °C µl	Error µl	Measurement Expanded Uncertainty ± µl
1.	Volume	5	4.8817	4.9017	-0.0983	0.039
2.	5 µl to 50 µl	20	19.7520	19.8328	-0.1672	0.39
3.		50	49.5560	49.7587	-0.2413	0.39

Remarks: i) Cubical Expansion co-efficient of pipette material taken as 10⁻⁵ µl / °k.

ii) This result has an expanded uncertainty with a coverage factor k=2 at approximately 95% confidence level.

iii) The calibration certificate issued for this instrument is to be used for scientific or industrial purposes only.

DUC - Device Under Calibration N.S. - Not Specified N.A. - Not Applicable

Opinions and Interpretations	
Calibrated	√ Accepted / Valid for use
Limited Use	Rejected / Out of use

Calibrated by:

K. Barat
K. Barat

Testing Engineer
Testing Engineer

Measure Techno Lab
Kolkata

Checked / Approved by:

SKP
Quality & Technical Manager

S. Pandey